



LABORATORY TESTING SERVICES

Soil, Water, Fuel & Waste



DETS is now SUEZ

CONTENTS

| | |
|---|-----------|
| Introduction..... | 4 |
| Market leading laboratory | 4 |
| Deviating Samples..... | 5 |
| Soils Summary | 6 |
| Notes | 10 |
| Waters Summary | 11 |
| Notes | 13 |
| Gases | 14 |
| Bulk Gas | 14 |
| C1 - C7 | 14 |
| Fuels | 15 |
| Sample preparation | 15 |
| Metals..... | 16 |
| Visual Inspection | 17 |
| PSD | 17 |
| Technical Information..... | 18 |
| Waste | 20 |
| Waste Characterisation and Leachate Analysis..... | 20 |
| INERT WAC | 20 |
| HAZARDOUS WAC | 21 |
| FULL WAC..... | 22 |
| Hazardous Waste / WM3POP Suite | 23 |
| Metals | 24 |
| Soils | 24 |
| Waters..... | 26 |
| Physical tests and wet chemistry..... | 27 |
| Soils | 27 |
| BARGE Bioaccessability Test (on soils) | 27 |
| Waters | 28 |
| Nitrogen compounds in water | 28 |
| Anions | 29 |
| Soils..... | 29 |
| Waters..... | 29 |

| | |
|--|-----------|
| Asbestos Analyses | 30 |
| Total Petroleum Hydrocarbons (TPH) | 31 |
| TPH Screen (soils only)..... | 31 |
| Volatile Petroleum Hydrocarbons (VPH) | 31 |
| Extractable petroleum hydrocarbons (EPH) | 31 |
| Speciated TPH (Criteria Working Group - CWG) | 32 |
| Volatile Organic Compounds (VOCs)..... | 33 |
| Semi Volatile Organic Compounds (SVOCs) | 37 |
| Polyaromatic Hydrocarbons | 39 |
| Polychlorinated Biphenyls (PCBs) | 41 |
| PCB 7 Congeners..... | 41 |
| PCB WHO 12 congeners (dioxin like PCBs)..... | 41 |
| PCB 7 congeners and PCB WHO 12 congeners (dioxin like) | 42 |
| Phenols..... | 43 |
| Phenols by GCMS | 43 |
| Pesticides | 44 |
| Combined Pesticide Suite | 44 |
| Organochlorine Pesticide Suite | 45 |
| Organophosphorous Pesticide Suite..... | 46 |
| Acid Herbicides | 47 |
| National Grid Suites | 48 |
| Monitored Natural Attenuation (MNA)..... | 51 |
| MNA Suite | 51 |
| Landfill Directive Lists..... | 52 |
| Landfill Directive List I | 52 |
| Landfill Directive List II | 52 |
| List 1 and List 2 substances | 53 |
| EA LIST 1 DANGEROUS SUBSTANCES (76/464/EEC) | 53 |
| EA LIST 1 DANGEROUS SUBSTANCES (76/464/EEC) | 53 |
| Environmental Forensic | 55 |
| Specialist Analysis..... | 56 |



Market leading laboratory

DETS is part of a global network of laboratories operating within the SUEZ group that provide a range of analytical testing and technical solutions to our customers.

With twenty years experience in the industry, we serve a client base of environmental consultants, site investigation companies, remediation companies, local authorities, regulators, waste companies and industrial manufacturers. We have two laboratories:

- Consett, Co. Durham (DETS North);
- Lenham, Kent (DETS South),

which are both accredited to ISO 17025 and MCERTS.

With a dedicated staff of 160+ and state of the art laboratories, we are well equipped to deal with varied and demanding client requirements. We also have our own team of drivers and vans, plus accreditation for sampling waters on site.

We offer a full range of environmental determinands including: Metals, organics (TPH, PAHs, VOCs, SVOCs, pesticides) all standard water quality parameters (pH, EC, anions, BOD, COD, TSS, TDS etc), plus specialist soil tests.

We also offer a range of specialist waste analyses including TOC, Lol, SRF, ADF, CV, WAC and other leaching tests.

DETS are recognised as a centre of excellence for asbestos analysis, and offer the following tests:

- Asbestos identification
- Asbestos quantification, including fibre dispersion and counting
- Respirable fibres
- Respirable fibres in respirable dust
- Classification for licensed or non-licensed work

This product catalogue provides a useful overview of our services. The information refers to our CONSETT laboratory. For specific details about our LENHAM laboratory please contact 01622850410 or enquiries@dets.co.uk.

The first section provides summary sheets for soils and waters, with the subsequent pages providing in depth information with respect to individual determinands, limits of detection, and accreditation status, plus a number of commonly requested suites and packages.



In light of the guidance provided by the Laboratory Committee of the European cooperation for Accreditation (EA) in document EA/LC (07) 60, UKAS would like to reiterate its position relating to the handling of deviating samples by accredited laboratories and the disclaimers required in test reports or certificates.

Deviating samples are samples which are not (correctly) preserved, for example they may have exceeded their maximum preservation time, lack the date and time of sampling, are not cooled, have inappropriate headspace and so on. As a result, deviating samples may jeopardize the validity of the reported test result.

Accreditation bodies which are members of EA have previously observed that laboratories, in particular those operating in highly competitive markets, were not critical about the samples they receive. Large numbers of deviating samples were accepted, analysed and test reports were issued without any remark. The EA Laboratory Committee members concluded that such practice is not in the interest of the laboratories, their customers or other end-users of the test result(s), nor of the accreditation bodies. Laboratories must therefore inform customers regarding samples which are deviating and include appropriate comments in their final reports.



v. 2.0 issued June 2020

Although every effort is made to ensure the accuracy of published information in this document, the information may inadvertently contain inaccuracies or typographical errors. Every effort has been made to present you with the most accurate information of the testing we provide, but because the nature of scientific and business research is constantly evolving, we cannot be held responsible for the accuracy of our content. DETS and SUEZ will not be held liable and shall not accept any liability for the content or interpretation of this document and reserve the right to change the document with no prior notice.

| Analyte | Minimum required (g) | Container G = 1L glass P = 1L plastic | Stability | LOD | Units | Preparation Method | Extraction Method | Analysis Method | Accreditation (MCERTS includes 17025) |
|-------------------------|----------------------|---|-----------|-------|-------|--------------------|------------------------|-----------------|---------------------------------------|
| Acid Herbicides | 250 | G | 2 weeks | 0.1 | mg/kg | As Received | Pentane/hexane/ether | LCMS MS | None |
| Alkalinity | 500 | G or P | 1 week | 10 | mg/kg | Dried & Crushed | Aqueous Extraction | Titration | None |
| Aluminium | 500 | G or P | 6 months | 1 | mg/kg | Dried & Crushed | Aqua Regia Digest | ICP OES | None |
| Ammoniacal Nitrogen | 500 | G or P | 3 days | 0.5 | mg/kg | As Received | Aqueous Extraction | Colorimetric | MCERTS |
| Antimony | 500 | G or P | 6 months | 1 | mg/kg | Dried & Crushed | Aqua Regia Digest | ICP OES | None |
| Arsenic | 500 | G or P | 6 months | 0.2 | mg/kg | Dried & Crushed | Aqua Regia Digest | ICP OES | MCERTS |
| Asbestos ID | 1000 | G or P | None | N/A | N/A | Mechanical | N/A | PL Microscope | 17025 |
| Asbestos Quantification | 1000 | G or P | None | 0.001 | % | Dispersion | N/A | PCO Microscope | 17025 |
| Asbestos Bulk ID | 500 | G or P | None | N/A | N/A | Mechanical | N/A | PL Microscope | 17025 |
| Barium | 500 | G or P | 6 months | 1.5 | mg/kg | Dried & Crushed | Aqua Regia Digest | ICP OES | MCERTS |
| Beryllium | 500 | G or P | 6 months | 0.2 | mg/kg | Dried & Crushed | Aqua Regia Digest | ICP OES | MCERTS |
| Boron, water soluble | 500 | G or P | 6 months | 0.2 | mg/kg | Dried & Crushed | Hot Aqueous Extraction | ICP OES | MCERTS |
| Bromide | 500 | G or P | 1 month | 1 | mg/kg | Dried & Crushed | Aqueous Extraction | IC | None |
| BTEX | 60 | 60ml glass | 2 weeks | 0.01 | mg/kg | As Received | Headspace | GC FID | MCERTS |
| Cadmium | 500 | G or P | 6 months | 0.1 | mg/kg | Dried & Crushed | Aqua Regia Digest | ICP OES | MCERTS |
| Calcium, available | 500 | G or P | 6 months | 0.1 | mg/kg | Dried & Crushed | Ammonium Nitrate | ICP OES | None |
| Carbonate | 500 | G or P | 4 weeks | 1 | % | Dried & Crushed | 1M Hydrochloric Acid | Titration | 17025 |
| Chloride water soluble | 500 | G or P | 1 month | 1 | mg/kg | Dried & Crushed | Aqueous Extraction | IC | 17025 |
| Chloride water soluble | 500 | G or P | 1 month | 0.01 | % | Dried & Crushed | Aqueous Extraction | Titration | 17025 |
| Chloride acid soluble | 500 | G or P | 1 month | 0.01 | % | Dried & Crushed | Nitric Acid Extraction | Titration | 17025 |
| Chromium | 500 | G or P | 6 months | 0.15 | mg/kg | Dried & Crushed | Aqua Regia Digest | ICP OES | MCERTS |
| Chromium Hexavalent | 500 | G or P | 1 month | 1 | mg/kg | As Received | Aqueous Extraction | Colorimetric | None |
| Cobalt | 500 | G or P | 6 months | 0.7 | mg/kg | Dried & Crushed | Aqua Regia Digest | ICP OES | MCERTS |
| Copper | 500 | G or P | 6 months | 0.2 | mg/kg | Dried & Crushed | Aqua Regia Digest | ICP OES | MCERTS |
| Copper, available | 500 | G or P | 6 months | 0.002 | mg/kg | Dried & Crushed | EDTA | ICP OES | None |

Note: This information is for our CONSETT laboratory. For specific details for our LENHAM Laboratory please contact 01622850410 or enquiries@dets.co.uk.

Continued on next page . .

| Analyte | Minimum required (g) | Container G = 1L glass P = 1L plastic | Stability | LOD | Units | Preparation Method | Extraction Method | Analysis Method | Accreditation (MCERTS includes 17025) |
|--|----------------------|---|-----------|------|-------|--------------------|--------------------|-----------------|---------------------------------------|
| Cyanide Total | 500 | G or P | 2 weeks | 0.1 | mg/kg | As Received | Sodium Hydroxide | Continuous Flow | MCERTS |
| Cyanide Free (easily liberatable) | 500 | G or P | 2 weeks | 0.1 | mg/kg | As Received | Sodium Hydroxide | Continuous Flow | MCERTS |
| Cyanide Complex | 500 | G or P | 2 weeks | 0.2 | mg/kg | N/A | N/A | Calculation | None |
| Cyclohexane Ext. Matter - CEM | 500 | G | 2 weeks | 50 | mg/kg | As Received | Cyclohexane | Gravimetric | None |
| Electrical Conductivity | 500 | G or P | 1 week | 1 | µS/cm | Dried & Crushed | Aqueous Extraction | Potentiometric | 17025 |
| Electrical Conductivity | 500 | G or P | 1 week | 1 | µS/cm | Dried & Crushed | Calcium Sulphate | Potentiometric | None |
| EPH (extractable petroleum hydrocarbons) | 250 | G | 2 weeks | 10 | mg/kg | As Received | DCM Extraction | GC FID | MCERTS |
| Fluoride | 500 | G or P | 1 month | 1 | mg/kg | Dried & Crushed | Aqueous Extraction | IC | 17025 |
| Iron Total | 500 | G or P | 6 months | 25 | mg/kg | Dried & Crushed | Aqua Regia Digest | ICP OES | 17025 |
| Lead | 500 | G or P | 6 months | 0.3 | mg/kg | Dried & Crushed | Aqua Regia Digest | ICP OES | MCERTS |
| Lithium | 500 | G or P | 6 months | 1 | mg/kg | Dried & Crushed | Aqua Regia Digest | ICP OES | None |
| Loss On Ignition | 500 | G or P | 4 weeks | 0.01 | % | Dried & Crushed | N/A | Gravimetric | MCERTS |
| Magnesium | 500 | G or P | 6 months | 1 | mg/kg | Dried & Crushed | Aqua Regia Digest | ICP OES | None |
| Magnesium, available | 500 | G or P | 6 months | 0.1 | mg/kg | Dried & Crushed | Ammonium Nitrate | ICP OES | None |
| Manganese | 500 | G or P | 6 months | 20 | mg/kg | Dried & Crushed | Aqua Regia Digest | ICP OES | MCERTS |
| Mercury | 500 | G or P | 4 weeks | 0.05 | mg/kg | Dried & Crushed | Aqua Regia Digest | AFS | MCERTS |
| Mercury Elemental | 100 | G or P | 4 weeks | 0.6 | µg/kg | As Received | Argon purge | AFS | None |
| Mercury Organic | 100 | G or P | 4 weeks | 100 | µg/kg | Dried & Crushed | Acid Digest | AFS | None |
| Mercury Inorganic | 100 | G or P | 4 weeks | 100 | µg/kg | Dried & Crushed | Acid Digest | AFS | None |
| Molybdenum | 500 | G or P | 6 months | 0.4 | mg/kg | Dried & Crushed | Aqua Regia Digest | ICP OES | MCERTS |
| Nickel | 500 | G or P | 6 months | 1 | mg/kg | Dried & Crushed | Aqua Regia Digest | ICP OES | MCERTS |

Note: This information is for our CONSETT laboratory. For specific details for our LENHAM Laboratory please contact 01622850410 or enquiries@dets.co.uk.

Continued on next page . .

| Analyte | Minimum required (g) | Container G = 1L glass P = 1L plastic | Stability | LOD | Units | Preparation Method | Extraction Method | Analysis Method | Accreditation (MCERTS includes 17025) |
|-------------------------------------|----------------------|---|-----------|------|----------|--------------------|-----------------------|-----------------|---------------------------------------|
| Nitrate | 500 | G or P | 1 month | 1 | mg/kg | Dried & Crushed | Aqueous Extraction | IC | 17025 |
| Nitrate | 500 | G or P | 1 month | 1 | mg/kg | Dried & Crushed | Aqueous Extraction | Colorimetric | None |
| Nitrite | 500 | G or P | 1 month | 1 | mg/kg | Dried & Crushed | Aqueous Extraction | IC | 17025 |
| Nitrite | 500 | G or P | 1 month | 0.5 | mg/kg | Dried & Crushed | Aqueous Extraction | Colorimetric | None |
| Nitrogen Total | 500 | G or P | 1 month | 0.01 | % | Dried & Crushed | N/A | Calculation | None |
| Nitrogen Kjeldahl | 500 | G or P | 1 month | 0.01 | % | Dried & Crushed | Sulphuric Acid Digest | Titration | None |
| Nitrogen Total Oxidised | 500 | G or P | 1 month | 7 | mg/kg | Dried & Crushed | Aqueous Extraction | Colorimetric | None |
| OCPs (organo-chlorine pesticides) | 250 | G | 1 month | 0.1 | mg/kg | As Received | Hexane/acetone | GC MS | None |
| OPPs (organophosphorous pesticides) | 250 | G | 2 weeks | 0.1 | mg/kg | As Received | Hexane/acetone | GC MS | None |
| Organic Matter | 500 | G or P | 28 days | 0.1 | % | Dried & Crushed | Sulphuric Acid Digest | Titration | MCERTS |
| pH | 500 | G or P | 1 week | 1 | pH units | Dried & Crushed | Aqueous Extraction | Potentiometric | MCERTS |
| PAHs | 250 | G | 2 weeks | 0.1 | mg/kg | As Received | Hexane/acetone | GC MS | 17025 |
| PCBs | 250 | G | 1 month | 0.01 | mg/kg | As Received | Hexane/acetone | GC MS | MCERTS |
| Phenols Monohydric | 500 | G or P | 2 weeks | 0.3 | mg/kg | As Received | Sodium Hydroxide | Continuous Flow | MCERTS |
| Phenols Speciated | 250 | G | 2 weeks | 0.01 | mg/kg | As Received | DCM | GC MS | None |
| Phosphate | 500 | G or P | 1 month | 0.1 | mg/kg | Dried & Crushed | Aqueous Extraction | Colorimetric | None |
| Phosphorous | 500 | G or P | 6 months | 1 | mg/kg | Dried & Crushed | Aqua Regia Digest | ICP OES | None |
| Phosphorous Available | 500 | G or P | 6 months | 0.1 | mg/kg | Dried & Crushed | NaHCO ₃ | ICP OES | None |
| Potassium | 500 | G or P | 6 months | 1 | mg/kg | Dried & Crushed | Aqua Regia Digest | ICP OES | None |
| Potassium Available | 500 | G or P | 6 months | 0.1 | mg/kg | Dried & Crushed | Ammonium Nitrate | ICP OES | None |
| PRO/GRO | 60 | 60ml glass | 2 weeks | 0.01 | mg/kg | As Received | Headspace | GC FID | None |
| Selenium | 500 | G or P | 6 months | 0.5 | mg/kg | Dried & Crushed | Aqua Regia Digest | ICP OES | MCERTS |
| Silica | 500 | G or P | 6 months | 10 | mg/kg | Dried & Crushed | None | XRF | None |

Note: This information is for our CONSETT laboratory . For specific details for our LENHAM Laboratory please contact 01622850410 or enquiries@dets.co.uk.

Continued on next page..

Soils Summary

| Analyte | Minimum required (g) | Container G = 1L glass P = 1L plastic | Stability | LOD | Units | Preparation Method | Extraction Method | Analysis Method | Accreditation (MCERTS includes 17025) |
|----------------------------------|----------------------|---|-----------|-------|-------|--------------------|--------------------|-----------------|---------------------------------------|
| Silver | 500 | G or P | 6 months | 1 | mg/kg | Dried & Crushed | Aqua Regia Digest | ICP OES | None |
| Sodium | 500 | G or P | 6 months | 0.1 | mg/kg | Dried & Crushed | Aqua Regia Digest | ICP OES | None |
| Sodium, available | 500 | G or P | 6 months | 0.1 | mg/kg | Dried & Crushed | Ammonium Nitrate | ICP OES | None |
| Strontium | 500 | G or P | 6 months | 1 | mg/kg | Dried & Crushed | Aqua Regia Digest | ICP OES | None |
| Sulphate Water Soluble | 500 | G or P | 1 month | 10 | mg/l | Dried & Crushed | Aqueous Extraction | ICP OES | MCERTS |
| Sulphate Water Soluble | 500 | G or P | 1 month | 1 | mg/l | Dried & Crushed | Aqueous Extraction | IC | 17025 |
| Sulphate Total | 500 | G or P | 1 month | 0.01 | % | Dried & Crushed | Aqua Regia Digest | ICP OES | MCERTS |
| Sulphide | 500 | G or P | 1 week | 10 | mg/kg | As Received | Distillation | Titration | None |
| Sulphur Total | 500 | G or P | 1 week | 0.01 | % | Dried & Crushed | Aqua Regia Digest | ICP OES | 17025 |
| Sulphur Elemental | 250 | G or P | 1 week | 0.75 | mg/kg | As Received | DCM | HPLC | MCERTS |
| SVOCs | 250 | G | 2 weeks | 0.1 | mg/kg | As Received | Hexane/acetone | GC MS | 17025 |
| Tellurium | 500 | G or P | 6 months | 1 | mg/kg | Dried & Crushed | Aqua Regia Digest | ICP OES | None |
| Thallium | 500 | G or P | 6 months | 1 | mg/kg | Dried & Crushed | Aqua Regia Digest | ICP OES | None |
| Thiocyanate | 500 | G or P | 2 weeks | 0.6 | mg/kg | As Received | Sodium Hydroxide | Continuous Flow | MCERTS |
| Tin | 500 | G or P | 6 months | 1 | mg/kg | Dried & Crushed | Aqua Regia Digest | ICP OES | 17025 |
| Titanium | 500 | G or P | 6 months | 1 | mg/kg | Dried & Crushed | Aqua Regia Digest | ICP OES | None |
| TOC (total organic carbon) | 500 | G or P | 4 weeks | 0.5 | % | Dried & Crushed | None | Auto analyser | MCERTS |
| TEM (Toluene Extractable Matter) | 500 | G | 2 weeks | 40 | mg/kg | As Received | Toluene | Gravimetric | None |
| TPH CWG C10 to C40 | 250 | G | 2 weeks | 10 | mg/kg | As Received | Hexane/DCM | GC FID | MCERTS |
| Triazines | 250 | G | 2 weeks | 0.1 | mg/kg | As Received | Hexane/acetone | GC MS | None |
| Vanadium | 500 | G or P | 6 months | 0.8 | mg/kg | Dried & Crushed | Aqua Regia Digest | ICP OES | MCERTS |
| VOCs | 60 | 60ml glass | 1 week | 0.01 | mg/kg | As Received | Headspace | GC MS | 17025 |
| Zinc | 500 | G or P | 6 months | 1 | mg/kg | Dried & Crushed | Aqua Regia Digest | ICP OES | MCERTS |
| Zinc, available | 500 | G or P | 6 months | 0.002 | mg/kg | Dried & Crushed | EDTA Extraction | ICP OES | None |
| Zinc, equivalent | 500 | G or P | 6 months | 0.002 | mg/kg | N/A | N/A | Calculation | None |

Note: This information is for our CONSETT laboratory. For specific details for our LENHAM Laboratory please contact 01622850410 or enquiries@dets.co.uk.

Notes

In most cases, the analysis can be performed on a significantly reduced weight of sample, but to provide as representative a sample as possible, these weights are recommended.

All determinands are reported on a dry weight basis, therefore sufficient soil must be provided to allow a moisture content to be analysed for all as received analysis.

A separate tub/bag must be supplied if an asbestos analysis is required, and these must be double bagged for transportation.



| Analyte | Minimum required (ml) | Container G = 1L glass P = 1L plastic | Stability | LOD | Unit | Prep/Extraction Method | Method | Accreditation |
|-----------------------------------|-----------------------|---|-----------|-------|-------|------------------------|-----------------|------------------|
| Acid Herbicides | 250 | G | 7 days | 0.02 | ug/l | n/a | LCMS MS | 17025 |
| Acidity | 500 | G or P | 2 weeks | 10 | mg/l | n/a | Titration | None |
| Alkalinity | 500 | G or P | 2 weeks | 10 | mg/l | n/a | Titration | 17025 |
| Aluminium | 500 | G or P | 6 months | 10 | ug/l | n/a | ICP MS | 17025 |
| Ammonia | 500 | G or P | 21 days | 0.015 | mg/l | n/a | Calculation | None |
| Ammoniacal Nitrogen | 500 | G or P | 21 days | 0.015 | mg/l | n/a | Colorimetric | 17025 |
| Antimony | 500 | G or P | 6 months | 0.17 | ug/l | n/a | ICP MS | 17025 |
| Arsenic | 500 | G or P | 6 months | 0.16 | ug/l | n/a | ICP MS | 17025 |
| Barium | 500 | G or P | 6 months | 0.26 | ug/l | n/a | ICP MS | 17025 |
| Beryllium | 500 | G or P | 6 months | 0.1 | ug/l | n/a | ICP MS | None |
| Bismuth | 500 | G or P | 6 months | 1 | ug/l | n/a | ICP MS | None |
| BOD | 500 | G or P | 2 days | 1 | mg/l | 5 day ATU | Meter | 17025 |
| Boron | 500 | G or P | 6 months | 12 | ug/l | n/a | ICP MS | None |
| Bromide | 500 | G or P | 1 month | 0.1 | mg/l | n/a | IC | None |
| BTEX | 40 | 40ml glass | 1 week | 1 | ug/l | Headspace | GC FID | 17025 |
| Cadmium | 500 | G or P | 6 months | 0.03 | ug/l | n/a | ICP MS | 17025 |
| Calcium | 500 | G or P | 6 months | 0.1 | mg/l | n/a | ICP MS | 17025 |
| Chloride | 500 | G or P | 1 month | 0.1 | mg/l | n/a | IC | 17025 |
| Chlorine | 250 | G or P | 1 day | 0.01 | mg/l | n/a | Kit | None |
| Chromium | 500 | G or P | 6 months | 0.25 | ug/l | n/a | ICP MS | 17025 |
| Chromium Hexavalent | 500 | G or P | 4 days | 7 | ug/l | n/a | Colorimetric | 17025 |
| Cobalt | 500 | G or P | 6 months | 0.16 | ug/l | n/a | ICP MS | 17025 |
| COD | 150 | G or P | 6 months | 10 | mg/l | Dichromate digestion | Colorimetric | 17025/ MCERTS |
| Copper | 500 | G or P | 6 months | 0.4 | ug/l | n/a | ICP MS | 17025 |
| Cyanide Total | 500 | G or P | 14 days | 0.1 | ug/l | Sodium Hydroxide | Continuous Flow | 17025 |
| Cyanide Free | 500 | G or P | 14 days | 0.1 | ug/l | Sodium Hydroxide | Continuous Flow | 17025 |
| Cyanide Complex | 500 | G or P | 14 days | 0.1 | ug/l | n/a | Calculation | 17025 |
| Cyclohexane Ext . Matter - CEM | 500 | G | 1 month | 1 | mg/l | Cyclohexane | Gravimetric | 17025 |
| Dissolved organic carbon (DOC) | 250 | G or P | 28 days | 2 | mg/l | Filtration | Auto Analyser | 17025 |
| Dissolved oxygen | 250 | G or P | 2 days | 0.1 | mg/l | n/a | Meter | None |
| Electrical Conductivity | 500 | G or P | 24 hours | 1 | uS/cm | n/a | Potentiometric | 17025 |
| EPH | 250 | G or P | 4 days | 10 | ug/l | DCM Extraction | GC FID | 17025 |
| Fluoride | 500 | G or P | 28 days | 0.1 | mg/l | n/a | IC | None |
| Hardness | 250 | G or P | 6 months | 0.1 | mg/l | n/a | Calculation | 17025 |
| Iodide | 250 | G or P | 28 days | 0.1 | mg/l | n/a | IC | None |
| Iron Total | 500 | G or P | 6 months | 5.5 | ug/l | n/a | ICP MS | 17025 |

Note: This information is for our CONSETT laboratory. For specific details for our LENHAM Laboratory please contact 01622850410 or enquiries@dets.co.uk.

Continued on next page..

| Analyte | Minimum required (ml) | Container G = 1L glass P = 1L plastic | Stability | LOD | Unit | Prep/Extraction Method | Method | Accreditation |
|-------------------------|-----------------------|---|-----------|-------|----------|------------------------|-----------------|---------------|
| Iron Ferric | 500 | G or P | 1 week | 0.1 | mg/l | n/a | Calculation | None |
| Iron Ferrous | 500 | G or P | 1 week | 0.1 | mg/l | Sulphuric Acid Digest | Colorimetric | None |
| Lead | 500 | G or P | 6 months | 0.09 | ug/l | n/a | ICP MS | 17025 |
| Lithium | 500 | G or P | 6 months | 1 | ug/l | n/a | ICP MS | None |
| Magnesium | 500 | G or P | 6 months | 0.02 | mg/l | n/a | ICP MS | 17025 |
| Manganese | 500 | G or P | 6 months | 0.22 | ug/l | n/a | ICP MS | 17025 |
| Mercury | 500 | G or P | 6 months | 0.001 | ug/l | n/a | AFS | 17025 |
| Mercury Elemental | 100 | G or P | 2 weeks | 1 | ug/l | Argon purge | AFS | None |
| Mercury Organic | 100 | G or P | 2 weeks | 1 | ug/l | n/a | AFS | None |
| Mercury Inorganic | 100 | G or P | 28 days | 1 | ug/l | n/a | AFS | None |
| Molybdenum | 500 | G or P | 6 months | 1.1 | ug/l | n/a | ICP MS | None |
| Nickel | 500 | G or P | 6 months | 0.5 | ug/l | n/a | ICP MS | 17025 |
| Nitrate | 500 | G or P | 28 days | 0.1 | mg/l | n/a | IC | 17025 |
| Nitrate | 500 | G or P | 28 days | 0.1 | mg/l | n/a | Colorimetric | None |
| Nitrite | 500 | G or P | 5 days | 0.1 | mg/l | n/a | IC | 17025 |
| Nitrite | 500 | G or P | 5 days | 0.035 | mg/l | n/a | Colorimetric | 17025 |
| Nitrogen Total | 500 | G or P | 1 month | 1 | mg/l | n/a | Calculation | None |
| Nitrogen Kjeldahl | 500 | G or P | 1 month | 0.2 | mg/l | Sulphuric Acid Digest | Titration | None |
| Nitrogen Total Oxidised | 500 | G or P | 5 days | 0.7 | mg/l | n/a | Colorimetric | 17025 |
| OCPs | 250 | G | 1 week | 1 | ug/l | DCM extraction | GC MS | None |
| OPPs | 250 | G | 1 week | 1 | ug/l | DCM extraction | GC MS | None |
| pH | 500 | G or P | 24 hours | 1 | pH units | n/a | Potentiometric | 17025 |
| PAHs (Total) | 250 | G | 4 days | 0.2 | ug/l | DCM extraction | GC MS | 17025 |
| PCBs (Total) | 250 | G | 1 week | 1 | ug/l | DCM extraction | GC MS | 17025 |
| Phenols Monohydric | 500 | G or P | 3 weeks | 1.5 | ug/l | Sodium Hydroxide | Continuous Flow | 17025 |
| Phenols Speciated | 250 | G | 3 weeks | 0.5 | ug/l | DCM | GC MS | None |
| Phosphate | 500 | G or P | 5 days | 0.01 | mg/l | n/a | Colorimetric | 17025 |
| Phosphorous | 500 | G or P | 6 months | 10 | ug/l | n/a | ICP MS | 17025 |
| Potassium | 500 | G or P | 6 months | 0.08 | mg/l | n/a | ICP MS | 17025 |
| PRO/GRO | 60 | 60ml glass | 7 days | 1 | ug/l | Headspace | GC FID | 17025 |
| Selenium | 500 | G or P | 6 months | 0.25 | ug/l | n/a | ICP MS | 17025 |
| Silver | 500 | G or P | 6 months | 0.13 | ug/l | n/a | ICP MS | None |
| Sodium | 500 | G or P | 6 months | 0.07 | mg/l | n/a | ICP MS | 17025 |
| Strontium | 500 | G or P | 6 months | 0.4 | ug/l | n/a | ICP MS | None |

Note: This information is for our CONSETT laboratory. For specific details for our LENHAM Laboratory please contact 01622850410 or enquiries@dets.co.uk.

| Analyte | Minimum required (ml) | Container G = 1L glass P = 1L plastic | Stability | LOD | Unit | Prep/Extraction Method | Method | Accreditation |
|------------------------------|-----------------------|---|-----------|--------|------|------------------------|---------------|---------------|
| Sulphate | 500 | G or P | 28 days | 0.01 | mg/l | n/a | ICP OES | None |
| Sulphate | 500 | G or P | 28 days | 0.1 | mg/l | n/a | IC | 17025 |
| Sulphide | 500 | G or P | 1 week | 10 | ug/l | n/a | Colorimetric | 17025 |
| Sulphur Total | 500 | G or P | 1 week | 10 | mg/l | n/a | ICP OES | None |
| Sulphur Elemental | 250 | G or P | 1 week | 84 | ug/l | DCM extraction | HPLC | 17025 |
| SVOCs | 250 | G | 1 week | 1 | ug/l | DCM extraction | GC MS | None |
| TDS | 250 | G or P | 1 week | 5 | mg/l | Filtration | Gravimetric | 17025 |
| Tellurium | 500 | G or P | 6 months | 0.1 | ug/l | n/a | ICP MS | None |
| Thallium | 500 | G or P | 6 months | 0.08 | ug/l | n/a | ICP MS | None |
| Thiocyanate | 500 | G or P | 3 days | 20 | ug/l | Sodium Hydroxide | Colorimetric | 17025 |
| Tin | 500 | G or P | 6 months | 0.4 | ug/l | n/a | ICP MS | None |
| Organo Tin compounds (total) | 500 | G | 2 weeks | 10 | ug/l | DCM Extraction | GCMS | None |
| Titanium | 500 | G or P | 6 months | 0.3 | ug/l | n/a | ICP MS | None |
| TOC | 500 | G or P | 28 days | 1 | mg/l | n/a | Auto analyser | 17025 |
| Toluene Ext . Matter - TEM | 500 | G | 1 month | 1 | mg/l | Toluene | Gravimetric | 17025 |
| TPH CWG C10 to C40 | 250 | G | 4 days | 10 | ug/l | Hexane/DCM | GC FID | None |
| Triazines | 250 | G | 1 month | 1 | ug/l | DCM extraction | GC MS | None |
| TSS | 250 | G or P | 48 hours | 5 | mg/l | Filtration | Gravimetric | 17025 |
| Vanadium | 500 | G or P | 6 months | 0.6 | ug/l | n/a | ICP MS | 17025 |
| VOCs | 40 | 40ml glass | 1 week | varies | ug/l | Headspace | GC MS | 17025 |
| Zinc | 500 | G or P | 6 months | 1.3 | ug/l | n/a | ICP MS | 17025 |

Note: This information is for our CONSETT laboratory . For specific details for our LENHAM Laboratory please contact 01622850410 or enquiries@dets.co.uk.

Notes

Most inorganic tests can be performed on a significantly reduced volume, but not the organics. If very low detection limits are required for organics, a larger volume may be required.

If preserved bottles are used, the samples must be filtered on site, particularly for metals.

Samples should be stored at 4.5 +/- 3.5°C as stated in the MCERTS standard for water testing.

Free ammonia is derived by calculation from the ammoniacal nitrogen and the pH of the sample.

BOD samples should always be taken in a separate bottle with no headspace.

Soil gas analysis is useful in determining if methane is generated from a site, or to provide a profile of possible contamination such as volatile solvents or petroleum products. Gases are usually sampled in Tedlar bags.

Bulk Gas

| Determinand | LOD | Units | Accreditation |
|-------------------|-------|-------|---------------|
| Methane | 0.005 | % | None |
| Carbon Dioxide | 0.005 | % | None |
| Oxygen | 0.005 | % | None |
| Nitrogen | 0.005 | % | None |
| Carbon Monoxide | 0.005 | % | None |
| Hydrogen | 1 | % | None |
| Hydrogen sulphide | 0.01 | % | None |

Note: This information is for our CONSETT laboratory. For specific details for our LENHAM Laboratory please contact 01622850410 or enquiries@dets.co.uk.

C1 - C7

| Determinand | LOD | Units | Accreditation |
|------------------|-------|-------|---------------|
| Acetylene | 0.005 | % | None |
| Ethylene | 0.005 | % | None |
| Ethane | 0.005 | % | None |
| Methyl acetylene | 0.005 | % | None |
| Propylene | 0.005 | % | None |
| Propane | 0.005 | % | None |
| Butane | 0.005 | % | None |

Note: This information is for our CONSETT laboratory. For specific details for our LENHAM Laboratory please contact 01622850410 or enquiries@dets.co.uk.



DETS offer a wide range of testing on recovered waste materials including SRF, ADF, compositional and elemental analysis .

| Test | LOD | Units | SRF - In house method based on: | SOLID BIOFUEL -In house method based on: |
|---|-------|-----------------------|---------------------------------|--|
| Sample preparation | | | | |
| Analysis Moisture | 0.10 | % | BS EN 15414 | BS EN 14774 |
| Total Moisture | 0.10 | % | BS EN 15414 | BS EN 14774 |
| Ash Content -Coal (as analysed, as received, dry basis) | 0.10 | % | ***** | ***** |
| Ash Content -Solid Biomass & SRF (as analysed, as received, dry basis) | 0.10 | % | BS EN 15403 | BS EN 14775 |
| Gross Calorific Value (as analysed, as received, dry & dry ash free basis) | 1.00 | MJ/kg | BS EN 15400 | BS EN 14918 |
| Net Calorific Value (as received & dry basis) | 1.00 | MJ/kg | BS EN 15400 | BS EN 14918 |
| Bulk Density | 0.50 | kg/m ³ | BS EN 15401 | BS EN 15103 |
| Biomass | 5 | % | BS EN 15403 | ***** |
| Non-biomass content (dry basis) | 5 | % | BS EN 15403 | ***** |
| Bromine | 0.05 | % | ***** | ***** |
| Chlorine (as analysed, as received, dry & dry ash free basis) | 0.05 | % | ***** | ***** |
| Fluorine | 0.035 | % | ***** | ***** |
| Iodine | 0.010 | % | ***** | ***** |
| Total Halides | 1.00 | % | ***** | ***** |
| Carbon (as analysed, as received, dry & dry ash free basis) | 0.1 | % | BS EN 15407 | BS EN 15104 |
| Hydrogen (as analysed, as received, dry & dry ash free basis) | 0.3 | % | BS EN 15407 | BS EN 15104 |
| Oxygen - by difference (as analysed, as received, dry & dry ash free basis) | 4 | % | ***** | ***** |
| Nitrogen (as analysed, as received, dry & dry ash free basis) | 0.3 | % | BS EN 15407 | BS EN 15104 |
| Sulphur (as analysed, as received, dry & dry ash free basis) | 0.010 | % | ***** | ***** |
| Fixed Carbon (as analysed, as received, dry & dry ash free basis) | 0.10 | % | ***** | ***** |
| Silicon | 10.0 | mg/kg | ***** | ***** |
| Cyanide total | 0.10 | mg/kg | ***** | ***** |
| Sulphide | 10.0 | mg/kg | ***** | ***** |
| Biomethane Potential | 0.10 | M ³ /Tonne | ***** | ***** |
| Volatile Matter (as analysed, as received, dry & dry ash free basis)) | 0.10 | % | BS EN 15402 | BS EN 15148 |

Note: This information is for our CONSETT laboratory . For specific details for our LENHAM Laboratory please contact 01622850410 or enquiries@dets.co.uk.



| Test | LOD | Units | SRF - In house method based on: | SOLID BIOFUEL - In house method based on: |
|---------------|-------|-------|---------------------------------|---|
| Metals | | | | |
| Aluminium | 1.00 | mg/kg | BS EN 15410 | BS EN 15290 |
| Antimony | 0.10 | mg/kg | BS EN 15411 | BS EN 15297 |
| Arsenic | 0.10 | mg/kg | BS EN 15411 | BS EN 15297 |
| Barium | 10.00 | mg/kg | BS EN 15411 | ***** |
| Beryllium | 0.10 | mg/kg | BS EN 15411 | ***** |
| Cadmium | 0.10 | mg/kg | BS EN 15411 | BS EN 15297 |
| Calcium | 5.00 | mg/kg | BS EN 15410 | BS EN 15290 |
| Chromium | 0.20 | mg/kg | BS EN 15411 | BS EN 15297 |
| Cobalt | 0.10 | mg/kg | BS EN 15411 | BS EN 15297 |
| Copper | 0.20 | mg/kg | BS EN 15411 | BS EN 15297 |
| Iron | 1.00 | mg/kg | BS EN 15410 | BS EN 15290 |
| Lead | 0.10 | mg/kg | BS EN 15411 | BS EN 15297 |
| Magnesium | 1.00 | mg/kg | BS EN 15410 | BS EN 15290 |
| Manganese | 0.10 | mg/kg | BS EN 15411 | BS EN 15297 |
| Mercury | 0.30 | mg/kg | BS EN 15411 | BS EN 15297 |
| Molybdenum | 0.50 | mg/kg | BS EN 15411 | BS EN 15297 |
| Nickel | 0.10 | mg/kg | BS EN 15411 | BS EN 15297 |
| Phosphorus | 0.10 | mg/kg | BS EN 15410 | BS EN 15290 |
| Potassium | 1.00 | mg/kg | BS EN 15410 | BS EN 15290 |
| Rhodium | 10.00 | mg/kg | ***** | ***** |
| Selenium | 0.10 | mg/kg | BS EN 15411 | ***** |
| Silver | 10.00 | mg/kg | ***** | ***** |
| Sodium | 5.00 | mg/kg | BS EN 15410 | BS EN 15290 |
| Strontium | 10.00 | mg/kg | ***** | ***** |
| Thallium | 0.20 | mg/kg | BS EN 15411 | ***** |
| Tin | 0.10 | mg/kg | ***** | ***** |
| Titanium | 0.10 | mg/kg | BS EN 15410 | BS EN 15290 |
| Vanadium | 0.10 | mg/kg | BS EN 15411 | BS EN 15297 |
| Zinc | 0.10 | mg/kg | BS EN 15411 | BS EN 15297 |

Note: This information is for our CONSETT laboratory . For specific details for our LENHAM Laboratory please contact 01622850410 or enquiries@dets.co.uk.



| Test | LOD | Units | SRF - In house method based on: | SOLID BIOFUEL - In house method based on: |
|--|------|-------|---------------------------------|---|
| Visual Inspection | | | | |
| Colour | n/a | n/a | ***** | ***** |
| Cellulosic Materials | 0.10 | % | ***** | ***** |
| Cloth Materials | 0.10 | % | ***** | ***** |
| Inert Materials | 0.10 | % | ***** | ***** |
| Metal Materials | 0.10 | % | ***** | ***** |
| Paper Materials | 0.10 | % | ***** | ***** |
| Plastics Materials | 0.10 | % | ***** | ***** |
| Biodegradable Municipal Waste Content (Compositional Analysis) | 0.10 | % | ***** | ***** |
| PSD | | | | |
| Passed through 2mm | 0.10 | % | BS EN 15415 | BS EN 15149 |
| Passed through 4mm | 0.10 | % | BS EN 15415 | BS EN 15149 |
| Retained on 16mm | 0.10 | % | BS EN 15415 | BS EN 15149 |
| Retained on 2mm | 0.10 | % | BS EN 15415 | BS EN 15149 |
| Retained on 30mm | 0.10 | % | BS EN 15415 | BS EN 15149 |
| Retained on 4mm | 0.10 | % | BS EN 15415 | BS EN 15149 |
| Retained on 8mm | 0.10 | % | BS EN 15415 | BS EN 15149 |

Note: This information is for our CONSETT laboratory. For specific details for our LENHAM Laboratory please contact 01622850410 or enquiries@dets.co.uk.



Technical Information

| Description | Det/Suite | Preparation | LOD | Units | Accreditation |
|---------------------------|------------|-------------|-------|-------|---------------|
| - Moisture, Total | MOIS-TOT-F | As Received | 0.1 | % | 17025 |
| - Antimony | Sb-oe3F | Air Dried | 0.1 | mg/kg | 17025 |
| - Arsenic | As-oe3F | Air Dried | 0.1 | mg/kg | 17025 |
| - Cadmium | Cd-oe3F | Air Dried | 0.1 | mg/kg | 17025 |
| - Chromium | Cr-oe3F | Air Dried | 0.2 | mg/kg | 17025 |
| - Cobalt | Co-oe3F | Air Dried | 0.1 | mg/kg | 17025 |
| - Copper | Cu-oe3F | Air Dried | 0.2 | mg/kg | 17025 |
| - Lead | Pb-oe3F | Air Dried | 0.1 | mg/kg | 17025 |
| - Manganese | Mn-oe3F | Air Dried | 0.1 | mg/kg | 17025 |
| - Mercury | Hg-ps3F | Air Dried | 0.3 | mg/kg | 17025 |
| - Nickel | Ni-oe3F | Air Dried | 0.1 | mg/kg | 17025 |
| - Selenium | Se-oe3F | Air Dried | 0.1 | mg/kg | 17025 |
| - Tellurium | Te-oe3F | Air Dried | 10 | mg/kg | None |
| - Thallium | Tl-oe3F | Air Dried | 0.2 | mg/kg | 17025 |
| - Tin | Sn-oe3F | Air Dried | 0.1 | mg/kg | 17025 |
| - Vanadium | V-oe3F | Air Dried | 0.1 | mg/kg | 17025 |
| - Zinc | Zn-oe3F | Air Dried | 0.1 | mg/kg | 17025 |
| - Bromine | Br2-an1F | Air Dried | 0.05 | % | None |
| - Chlorine (as received) | Cl2-AR1F | As Received | 0.05 | % | 17025 |
| - Chlorine (dry basis) | Cl2-DRY1F | Air Dried | 0.05 | % | 17025 |
| - Chlorine (as analysed) | Cl2-an1F | Air Dried | 0.05 | % | 17025 |
| - Chlorine (dry ash free) | Cl2-DAF1F | Air Dried | 0.05 | % | 17025 |
| - Fluorine | Fl2-an1F | Air Dried | 0.035 | % | 17025 |
| - Iodine | Io-an1F | Air Dried | 0.01 | % | None |
| - Halides, Total | HAL-TOT1F | Air Dried | 1 | % | None |

Note: This information is for our CONSETT laboratory. For specific details for our LENHAM Laboratory please contact 01622850410 or enquiries@dets.co.uk.



Continued on next page . .

| Description | Det/Suite | Preparation | LOD | Units | Accreditation |
|--|------------|--------------|------|-------|---------------|
| - Gross Calorific Value (as received) | GCV-REC-F | As Received | 1 | MJ/kg | 17025 |
| - Gross Calorific Value (dry basis) | GCV-DRY-F | Dry basis | 1 | MJ/kg | 17025 |
| - Gross Calorific Value (dry ash free) | GCV-DAF-F | Dry ash free | 1 | MJ/kg | 17025 |
| - Net Calorific Value (as received) | NCV-REC-F | As Received | 1 | MJ/kg | 17025 |
| - Net Calorific Value (dry basis) | NCV-DRY-F | Dry basis | 1 | MJ/kg | 17025 |
| - Ash Content (dry basis) | ASH-DRY-F | Dry basis | 0.1 | % | 17025 |
| - Carbon (as received) | CARB-REC-F | As Received | 0.1 | % | 17025 |
| - Carbon (dry basis) | CARB-DRY-F | Dry basis | 0.1 | % | 17025 |
| - Carbon (dry ash free) | CARB-DAF-F | Dry ash free | 0.1 | % | 17025 |
| - Sulphur (as received) | SUL-REC-F | As Received | 0.01 | % | 17025 |
| - Sulphur (dry basis) | SUL-DRY-F | Dry basis | 0.01 | % | 17025 |
| - Sulphur (dry ash free) | SUL-DAF-F | Dry ash free | 0.01 | % | 17025 |
| - Biomass (dry basis) by Calorific Value | BIOMASS2-F | Dry basis | 1 | % | 17025 |
| - Ferrous Metal | FERR-F | As Received | 0.1 | % | None |
| - Non-Ferrous Metal | NONFERR-F | As Received | 0.1 | % | None |
| - Glass | GLASS-F | As Received | 0.1 | % | None |
| - Organic Materials | ORG-F | As Received | 0.1 | % | None |
| - Carpet & Mats | C&M-F | As Received | 0.1 | % | None |
| - Fines <10mm | FINE10-F | As Received | 0.1 | % | None |
| - Leather & Rubber | L&R-F | As Received | 0.1 | % | None |
| - Paper & Cardboard | P&C-F | As Received | 0.1 | % | None |
| - Rigid Plastic | RIGIDP-F | As Received | 0.1 | % | None |
| - Soft Plastic | SOFTP-F | As Received | 0.1 | % | None |
| - Textiles | TEXTILES-F | As Received | 0.1 | % | None |
| - Tissue | TISS-F | As Received | 0.1 | % | None |
| - Stones | STONES-F | As Received | 0.1 | % | None |
| - Wood | WOOD-F | As Received | 0.1 | % | None |

Note: This information is for our CONSETT laboratory . For specific details for our LENHAM Laboratory please contact 01622850410 or enquiries@dets.co.uk .



Waste Characterisation and Leachate Analysis

Classification of a waste requires both characterisation of the solid material and also a leaching test to determine the possibility of mobile contaminants contaminating the surrounding environment. There are various leaching tests available, but the most commonly requested is based on BS EN 12457 - 2. The solid material is agitated for 24 hours, filtered, and the resulting eluate analysed for a range of determinands.



INERT WAC

| Determinand | LOD | Units | Accreditation for analysis only, not for leachate preparation |
|---|------|-------|---|
| Eluates for compliance using BS EN 12457 - 2 (1 batch) | | | |
| Arsenic | 0.16 | ug/l | 17025 |
| Barium | 0.26 | ug/l | 17025 |
| Cadmium | 0.03 | ug/l | 17025 |
| Chromium | 0.25 | ug/l | 17025 |
| Copper | 0.40 | ug/l | 17025 |
| Mercury | 0.01 | ug/l | 17025 |
| Molybdenum | 1.10 | ug/l | None |
| Nickel | 0.50 | ug/l | 17025 |
| Lead | 0.09 | ug/l | 17025 |
| Antimony | 0.17 | ug/l | 17025 |
| Selenium | 0.25 | ug/l | 17025 |
| Zinc | 1.30 | ug/l | 17025 |
| Chloride | 0.10 | mg/l | 17025 |
| Fluoride | 0.10 | mg/l | None |
| Sulphate | 0.10 | mg/l | 17025 |
| Total Dissolved Solids (TDS) | 5 | mg/l | None |
| Phenol Index | 100 | ug/l | 17025 |
| Dissolved Organic Carbon at own pH or pH 7.5-8.0 | 2 | mg/l | 17025 |
| Solid Suites | | | |
| Total Organic Carbon | 0.5 | % | MCERTS |
| BTEX | 0.04 | mg/kg | MCERTS |
| PCBs (7 congeners) | 0.01 | mg/kg | MCERTS |
| EPH C10-C40 | 10 | mg/kg | MCERTS |
| PAHs - total 17 including coronene | 1.6 | mg/kg | 17025 except coronene |

Note: This information is for our CONSETT laboratory. For specific details for our LENHAM Laboratory please contact 01622850410 or enquiries@dets.co.uk.

HAZARDOUS WAC

| Determinand | LOD | Units | Accreditation for analysis only |
|---|------|----------|---------------------------------|
| Eluates for compliance using BS EN 12457 - 2 (1 batch) | | | |
| Arsenic | 0.16 | ug/l | 17025 |
| Barium | 0.26 | ug/l | 17025 |
| Cadmium | 0.03 | ug/l | 17025 |
| Chromium | 0.25 | ug/l | 17025 |
| Copper | 0.40 | ug/l | 17025 |
| Mercury | 0.01 | ug/l | 17025 |
| Molybdenum | 1.10 | ug/l | None |
| Nickel | 0.50 | ug/l | 17025 |
| Lead | 0.09 | ug/l | 17025 |
| Antimony | 0.17 | ug/l | 17025 |
| Selenium | 0.25 | ug/l | 17025 |
| Zinc | 1.30 | ug/l | 17025 |
| Chloride | 0.10 | mg/l | 17025 |
| Fluoride | 0.10 | mg/l | None |
| Sulphate | 0.10 | mg/l | 17025 |
| Total Dissolved Solids (TDS) | 5 | mg/l | None |
| Phenol Index | 100 | ug/l | 17025 |
| Dissolved Organic Carbon at own pH or pH 7.5-8.0 | 2 | mg/l | 17025 |
| Solid Suites | | | |
| Total Organic Carbon | 0.5 | % | MCERTS |
| Loss on ignition | 0.01 | % | MCERTS |
| Acid neutralisation capacity | 1 | moles/kg | None |

Note: This information is for our CONSETT laboratory. For specific details for our LENHAM Laboratory please contact 01622850410 or enquiries@dets.co.uk.



FULL WAC

| Determinand | LOD | Units | Accreditation |
|---|------|----------|-----------------------|
| Eluates for compliance using BS EN 12457 - 2 (1 batch) | | | |
| Arsenic | 0.16 | ug/l | 17025 |
| Barium | 0.26 | ug/l | 17025 |
| Cadmium | 0.03 | ug/l | 17025 |
| Chromium | 0.25 | ug/l | 17025 |
| Copper | 0.40 | ug/l | 17025 |
| Mercury | 0.01 | ug/l | 17025 |
| Molybdenum | 1.10 | ug/l | None |
| Nickel | 0.50 | ug/l | 17025 |
| Lead | 0.09 | ug/l | 17025 |
| Antimony | 0.17 | ug/l | 17025 |
| Selenium | 0.25 | ug/l | 17025 |
| Zinc | 1.30 | ug/l | 17025 |
| Chloride | 0.10 | mg/l | 17025 |
| Fluoride | 0.10 | mg/l | None |
| Sulphate | 0.10 | mg/l | 17025 |
| Total Dissolved Solids (TDS) | 5 | mg/l | None |
| Phenol Index | 100 | ug/l | 17025 |
| Dissolved Organic Carbon at own pH or pH 7.5-8.0 | 2 | mg/l | 17025 |
| Solid Suites | | | |
| Total Organic Carbon | 0.5 | % | MCERTS |
| Loss on ignition | 0.01 | % | MCERTS |
| BTEX | 0.04 | mg/kg | MCERTS |
| PCBs (7 congeners) | 0.01 | mg/kg | MCERTS |
| EPH C10-C40 | 10 | mg/kg | MCERTS |
| pH | 1 | pH units | MCERTS |
| Acid neutralisation capacity | 1 | moles/kg | None |
| PAHs - total 17 including coronene | 1.6 | mg/kg | 17025 except coronene |

Note: This information is for our CONSETT laboratory. For specific details for our LENHAM Laboratory please contact 01622850410 or enquiries@dets.co.uk.



Hazardous Waste / WM3 POP Suite

| Determinand | LOD | Units | Accreditation |
|---|---------|-------|---------------|
| Dioxins and Furans : POPS Hazardous Waste / WM3 Suite | Various | ug/kg | Subcontracted |
| POPs basic GC-MSMS / LC-MSMS Suite | Various | mg/kg | Subcontracted |
| DDT (1,1,1-trichloro-2,2-bis (4-chlorophenyl)ethane) | 0.1 | mg/kg | None |
| Chlordane | 0.1 | mg/kg | None |
| Hexachlorocyclohexanes, including lindane | 0.1 | mg/kg | None |
| Dieldrin | 0.1 | mg/kg | None |
| Endrin | 0.1 | mg/kg | None |
| Heptachlor | 0.1 | mg/kg | None |
| Hexachlorobenzene | 0.1 | mg/kg | None |
| Chlordecone | 0.1 | mg/kg | None |
| Aldrin | 0.1 | mg/kg | None |
| Pentachlorobenzene | 0.1 | mg/kg | None |
| Mirex | 0.1 | mg/kg | None |
| Hexabromobiphenyl | 0.1 | mg/kg | None |
| Toxaphene species by GCMS | 0.1 | mg/kg | None |
| Total PCBs (aroclor 1254 or 1260) | 0.01 | mg/kg | None |
| Flamability | N/A | N/A | None |

Note: This information is for our CONSETT laboratory. For specific details for our LENHAM Laboratory please contact 01622850410 or enquiries@dets.co.uk.



Metals are determined using ICP-OES or ICP MS, and the dried and crushed soils are digested in aqua regia prior to analysis. Waters are filtered and then acidified. We can determine metals on soils using XRF, but this will give a total measurement, expressed as the oxide, which may be higher than by ICP.

Soils

Metals on soils are performed on a dried and crushed sample, apart from speciated mercury.

| | Method | LOD | Units | Accreditation |
|--------------------|---------------|-------|-------|---------------|
| Aluminium | ICP OES | 1 | mg/kg | None |
| Antimony | ICP OES | 1 | mg/kg | None |
| Arsenic | ICP OES | 0.2 | mg/kg | MCERTS |
| Barium | ICP OES | 1.5 | mg/kg | MCERTS |
| Beryllium | ICP OES | 0.2 | mg/kg | MCERTS |
| Boron (WS) | ICP OES | 0.2 | mg/kg | MCERTS |
| Cadmium | ICP OES | 0.1 | mg/kg | MCERTS |
| Calcium | ICP OES | 1 | mg/kg | None |
| Calcium (avail.) | ICP OES | 0.1 | mg/kg | None |
| Chromium | ICP OES | 0.15 | mg/kg | MCERTS |
| Chromium (hex.) | Colourimetric | 1 | mg/kg | None |
| Cobalt | ICP OES | 0.7 | mg/kg | MCERTS |
| Copper | ICP OES | 0.2 | mg/kg | MCERTS |
| Copper (avail.) | ICP OES | 0.002 | mg/kg | None |
| Iron | ICP OES | 25 | mg/kg | 17025 |
| Lead | ICP OES | 0.3 | mg/kg | MCERTS |
| Lithium | ICP OES | 1 | mg/kg | None |
| Magnesium | ICP OES | 1 | mg/kg | None |
| Magnesium (avail.) | ICP OES | 0.1 | mg/kg | None |
| Manganese | ICP OES | 20 | mg/kg | MCERTS |

Note: This information is for our CONSETT laboratory. For specific details for our LENHAM Laboratory please contact 01622850410 or enquiries@dets.co.uk.



Continued on next page . .

| | Method | LOD | Units | Accreditation |
|--------------------------------|-------------|-------|-------|---------------|
| Mercury (total) | AFS | 0.05 | mg/kg | MCERTS |
| Mercury (elemental) | AFS | 0.6 | ug/kg | None |
| Mercury (organic) | AFS | 100 | ug/kg | None |
| Mercury (inorganic) | AFS | 100 | ug/kg | None |
| Molybdenum | ICP OES | 0.4 | mg/kg | MCERTS |
| Nickel | ICP OES | 1 | mg/kg | MCERTS |
| Phosphorus | ICP OES | 1 | mg/kg | None |
| Phosphorus (avail.) | ICP OES | 0.1 | mg/kg | None |
| Potassium | ICP OES | 1 | mg/kg | None |
| Potassium (avail.) | ICP OES | 0.1 | mg/kg | None |
| Selenium | ICP OES | 0.5 | mg/kg | MCERTS |
| Silicon (as SiO ₂) | XRF | 10 | mg/kg | None |
| Silver | ICP OES | 1 | mg/kg | None |
| Sodium | ICP OES | 0.1 | mg/kg | None |
| Sodium (avail.) | ICP OES | 0.1 | mg/kg | None |
| Strontium | ICP OES | 1 | mg/kg | None |
| Sulphur (total) | ICP OES | 0.01 | % | 17025 |
| Sulphur, elemental | HPLC | 0.75 | mg/kg | MCERTS |
| Tellurium | ICP OES | 1 | mg/kg | None |
| Thallium | ICP OES | 1 | mg/kg | None |
| Tin | ICP OES | 1 | mg/kg | 17025 |
| Titanium | ICP OES | 1 | mg/kg | None |
| Vanadium | ICP OES | 0.8 | mg/kg | MCERTS |
| Zinc | ICP OES | 1 | mg/kg | MCERTS |
| Zinc (avail.) | ICP OES | 0.002 | mg/kg | None |
| Zinc (equivalent) | Calculation | 0.002 | mg/kg | None |

Note: This information is for our CONSETT laboratory. For specific details for our LENHAM Laboratory please contact 01622850410 or enquiries@dets.co.uk.



Waters

Metals on waters can be performed on a filtered sample (dissolved metals), or an unfiltered sample (total metals).

| | Method | LOD | Units | Accreditation |
|--------------------------------|---------------|-----------|-------|---------------|
| | | Dissolved | | |
| Aluminium | ICP MS | 10 | ug/l | 17025 |
| Antimony | ICP MS | 0.17 | ug/l | 17025 |
| Arsenic | ICP MS | 0.16 | ug/l | 17025 |
| Barium | ICP MS | 0.26 | ug/l | 17025 |
| Beryllium | ICP MS | 0.10 | ug/l | None |
| Bismuth | ICP MS | 1 | ug/l | None |
| Boron | ICP MS | 12 | ug/l | None |
| Cadmium | ICP MS | 0.03 | ug/l | 17025 |
| Calcium | ICP MS | 0.1 | mg/l | 17025 |
| Chromium | ICP MS | 0.25 | ug/l | 17025 |
| Chromium (hex .) | Colourimetric | 7 | ug/l | 17025 |
| Cobalt | ICP MS | 0.16 | ug/l | 17025 |
| Copper | ICP MS | 0.4 | ug/l | 17025 |
| Iron | ICP MS | 5.5 | ug/l | 17025 |
| Iron (ferric/ferrous) | Colourimetric | 0.1 | mg/l | None |
| Lead | ICP MS | 0.09 | ug/l | 17025 |
| Lithium | ICP MS | 1 | ug/l | None |
| Magnesium | ICP MS | 0.02 | mg/l | 17025 |
| Manganese | ICP MS | 0.22 | ug/l | 17025 |
| Mercury (total) | ICP MS | 0.01 | ug/l | 17025 |
| Mercury (elemental) | AFS | 1 | ug/l | None |
| Mercury (organic) | AFS | 1 | ug/l | None |
| Mercury (inorganic) | AFS | 1 | ug/l | None |
| Molybdenum | ICP MS | 1.1 | ug/l | None |
| Nickel | ICP MS | 0.5 | ug/l | 17025 |
| Phosphorus | ICP MS | 10 | ug/l | 17025 |
| Potassium | ICP MS | 0.08 | mg/l | 17025 |
| Selenium | ICP MS | 0.25 | ug/l | 17025 |
| Silicon (as SiO ₂) | Colourimetric | 0.1 | mg/l | None |
| Silver | ICP MS | 0.13 | ug/l | None |
| Sodium | ICP MS | 0.07 | mg/l | None |
| Strontium | ICP MS | 0.4 | ug/l | None |
| Sulphur (total) | ICP OES | 10 | mg/l | None |
| Sulphur, elemental | HPLC | 84 | ug/l | 17025 |
| Tellurium | ICP MS | 0.1 | ug/l | None |
| Thallium | ICP MS | 0.08 | ug/l | None |
| Tin | ICP MS | 0.4 | ug/l | None |
| Titanium | ICP MS | 0.3 | ug/l | None |
| Vanadium | ICP MS | 0.6 | ug/l | 17025 |
| Zinc | ICP MS | 1.3 | ug/l | 17025 |

Note: This information is for our CONSETT laboratory. For specific details for our LENHAM Laboratory please contact 01622850410 or enquiries@dets.co.uk.

This section includes physical tests, plus the standard water quality parameters.

Some parameters should be measured on site for optimum results, e.g. pH, dissolved oxygen, redox potential.

BOD is a time critical analysis and should be analysed within 24 hours of sampling, although the test takes 5 days. The sample should be taken in a separate 100 ml bottle, with no headspace.

Soils

| | Method | LOD | Units | Accreditation |
|----------------------------|----------------|------|----------|---------------|
| pH | Potentiometric | 1 | pH units | MCERTS |
| Conductivity | Potentiometric | 1 | uS/cm | 17025 |
| Loss on ignition | Gravimetric | 0.01 | % | MCERTS |
| Moisture content | Gravimetric | 0.01 | % | 17025 |
| Ash content | Gravimetric | 0.01 | % | None |
| Flashpoint | Closed cup | n/a | °C | None |
| Particle size distribution | Sieve | n/a | % | None |
| Stone content | Gravimetric | 0.1 | % | None |
| Calorific value | Calorimeter | 1 | Mj/kg | 17025 |

Note: This information is for our CONSETT laboratory. For specific details for our LENHAM Laboratory please contact 01622850410 or enquiries@dets.co.uk.

BARGE Bioaccessability Test (on soils)

This test is designed to mimic the uptake of contaminants from the gastro intestinal system. Soils are extracted in a series of reagents to facilitate this uptake, followed by analysis for a range of metals. This parameter can be extremely useful in assessing human health risk and can reduce expensive remediation.



Waters

| | Method | LOD | Units | Accreditation |
|---------------------------------|----------------|-----|-------------------|-----------------|
| pH | Potentiometric | 1 | pH units | 17025 |
| Conductivity | Potentiometric | 1 | uS/cm | 17025 |
| Specific gravity | Gravimetric | N/A | g/cm ³ | None |
| Redox potential | Meter | N/A | mV | None |
| Dissolved oxygen | Meter | 2 | mg/l | None |
| Turbidity | Meter | 1 | NTU | None |
| Total dissolved solids | Gravimetric | 5 | mg/l | 17025 |
| Total suspended solids | Gravimetric | 5 | mg/l | 17025 |
| Volatile suspended solids | Gravimetric | 5 | mg/l | None |
| Non-volatile suspended solids | Gravimetric | 5 | mg/l | None |
| BOD | 5 day ATU | 1 | mg/l | 17025 |
| COD | Colourimetric | 10 | mg/l | 17025 (MCERTS*) |
| Alkalinity as CaCO ₃ | Titration | 10 | mg/l | 17025 |
| Acidity | Titration | 10 | mg/l | None |

**Only for trade effluent to sewer*

Note: This information is for our CONSETT laboratory. For specific details for our LENHAM Laboratory please contact 01622850410 or enquiries@dets.co.uk.

Nitrogen compounds in water

| | Method | LoD | Units | Accreditation |
|---|---------------------|-------|-------|---------------|
| Ammoniacal nitrogen (NH ₃ + NH ₄) | Colourimetric | 0.015 | mg/l | 17025 |
| Free ammonia (NH ₃)* | Calculation | 0.015 | mg/l | None |
| Kjeldahl nitrogen (amm . N + org . N) | Digestion/titration | 0.2 | mg/l | None |
| Organic nitrogen (Kjeldahl N - amm . N) | Calculation | 1 | mg/l | None |
| Total nitrogen (NO ₃ + NO ₂ + Kjeldahl N) | Calculation | 1 | mg/l | None |
| Nitrate (NO ₃) | Colourimetric | 0.1 | mg/l | None |
| Nitrate (NO ₃) | IC | 0.1 | mg/l | 17025 |
| Nitrite (NO ₂) | Colourimetric | 0.035 | mg/l | 17025 |
| Nitrite (NO ₂) | IC | 0.1 | mg/l | 17025 |
| Total oxidised nitrogen (NO ₂ + NO ₃) | Colourimetric | 0.7 | mg/l | 17025 |

**Need pH for this*

Note: This information is for our CONSETT laboratory. For specific details for our LENHAM Laboratory please contact 01622850410 or enquiries@dets.co.uk.

Anions can be determined either by ion chromatography, colourimetric spectroscopy or by ICPI.

Water samples are filtered prior to analysis, and soils are usually extracted in water.

Soils

| | | Method | LOD | Units | Accreditation |
|---------------------------------|---------------|---------------|------|-------|---------------|
| Chloride | Water soluble | IC | 1 | mg/kg | 17025 |
| Sulphate | Water soluble | ICP OES | 1 | mg/kg | MCERTS |
| Sulphate | Acid soluble | ICP OES | 0.01 | % | MCERTS |
| Nitrate | Water soluble | IC | 1 | mg/kg | 17025 |
| Nitrite | Water soluble | IC | 1 | mg/kg | 17025 |
| Phosphate | Water soluble | Colourimetric | 0.1 | mg/kg | None |
| Fluoride | Water soluble | IC | 1 | mg/kg | 17025 |
| Bromide | Water soluble | IC | 1 | mg/kg | None |
| Sulphide | Acid soluble | Titration | 10 | mg/kg | None |
| Sulphur, total | | ICP OES | 0.01 | % | 17025 |
| Sulphur, elemental | | HPLC | 0.75 | mg/kg | MCERTS |
| Total cyanide | | Colourimetric | 0.1 | mg/kg | MCERTS |
| Free (easily liberated cyanide) | | Colourimetric | 0.1 | mg/kg | MCERTS |
| Complex cyanide | | Calculation | 0.2 | mg/kg | None |
| Thiocyanate | | Colourimetric | 0.6 | mg/kg | MCERTS |

Note: This information is for our CONSETT laboratory. For specific details for our LENHAM Laboratory please contact 01622850410 or enquiries@dets.co.uk.

Waters

| | | Method | LOD | Units | Accreditation |
|---------------------------------|--|---------------|-------|-------|---------------|
| Chloride | | IC | 0.1 | mg/l | 17025 |
| Sulphate | | IC | 0.1 | mg/l | 17025 |
| Nitrate | | Colourimetric | 0.1 | mg/l | None |
| Nitrate | | IC | 0.1 | mg/l | 17025 |
| Nitrite | | Colourimetric | 0.035 | mg/l | 17025 |
| Nitrite | | IC | 0.1 | mg/l | 17025 |
| Phosphate | | Colourimetric | 0.01 | mg/l | 17025 |
| Fluoride | | IC | 0.1 | mg/l | None |
| Bromide | | IC | 0.1 | mg/l | None |
| Sulphide | | Colourimetric | 10 | ug/l | 17025 |
| Sulphur, total | | ICP OES | 10 | mg/l | None |
| Total cyanide | | Colourimetric | 0.1 | ug/l | 17025 |
| Free (easily liberated cyanide) | | Colourimetric | 0.1 | ug/l | 17025 |
| Complex cyanide | | Calculation | 0.1 | ug/l | 17025 |
| Thiocyanate | | Colourimetric | 20 | ug/l | 17025 |

Note: This information is for our CONSETT laboratory. For specific details for our LENHAM Laboratory please contact 01622850410 or enquiries@dets.co.uk.

Samples for asbestos analysis should always be placed in a separate double bagged container. It is preferable to take approximately 1 kg of soil for analysis, due to the lack of homogeneity of asbestos.

| | | LOD | Accreditation |
|---------|---|--------|---------------|
| Stage 1 | Identification: presence/absence of asbestos fibres using Polarised Light Microscopy as per HSG248 | n/a | 17025 |
| Stage 2 | Quantification: detailed gravimetric, using PLM | 0.001% | 17025 |
| Stage 3 | Quantification: fibre dispersion, measurement and counting, using PLM and PCM | 0.001% | 17025 |
| | Respirable fibres - mass of asbestos fibres in range 1- 5 microns | 0.001% | None |
| | Potentially respirable fibres | 0.001% | None |
| | Respirable fibres in respirable dust - new test used in risk assessment for airborne fibres on site | 0.001% | None |
| | Classification into licensed or non-licensed work - moisture absorption | n/a | 17025 |

Note: This information is for our CONSETT laboratory. For specific details for our LENHAM Laboratory please contact 01622850410 or enquiries@dets.co.uk.



Total Petroleum Hydrocarbons (TPH)

TPH represents an extremely complex group of compounds (over 10,000), consisting of aliphatic, aromatic and NSO (nitrogen, sulphur, oxygen containing compounds).

Analysis is by GC/FID, with headspace for the volatile petroleum hydrocarbons (VPH C5 - 10), and DCM based solvent extraction for the heavier hydrocarbons (EPH C10 - 40).

VPH is often referred to as PRO (petrol range organics) or GRO (gasoline range organics).

VPH samples should be collected in 40ml glass vials for waters, and 60 g glass jars for soils, in duplicate to allow for repeats.

TPH Screen (soils only)

| | LOD | Units | Accreditation |
|-------------------------------|-----|-------|---------------|
| TPH (C6-40) single extraction | 10 | mg/kg | None |

Note: This information is for our CONSETT laboratory. For specific details for our LENHAM Laboratory please contact 01622850410 or enquiries@dets.co.uk.

Volatile Petroleum Hydrocarbons (VPH)

Soils

Waters

| | LOD | Units | Accreditation | LOD | Units | Accreditation |
|-------------------------|------|-------|---------------|-----|-------|---------------|
| Benzene | 0.01 | mg/kg | MCERTS | 1 | ug/l | 17025 |
| Toluene | 0.01 | mg/kg | MCERTS | 1 | ug/l | 17025 |
| Ethylbenzene | 0.01 | mg/kg | MCERTS | 1 | ug/l | 17025 |
| Xylenes (mixed isomers) | 0.01 | mg/kg | MCERTS | 1 | ug/l | 17025 |
| BTEX, total | 0.01 | mg/kg | MCERTS | 1 | ug/l | 17025 |
| C5 - 10, VPH | 0.01 | mg/kg | None | 1 | ug/l | 17025 |

Note: This information is for our CONSETT laboratory. For specific details for our LENHAM Laboratory please contact 01622850410 or enquiries@dets.co.uk.

Extractable petroleum hydrocarbons (EPH)

Soils

Waters

| | LOD | Units | Accreditation | LOD | Units | Accreditation |
|--|-----|-------|---------------|-----|-------|---------------|
| EPH, C10 - 40 | 10 | mg/kg | MCERTS | 10 | ug/l | 17025 |
| Cleaned up EPH (removes NSOs) | 10 | mg/kg | None | 10 | ug/l | None |
| Mineral oil (removes aromatics and NSOs) | 10 | mg/kg | None | 10 | ug/l | None |
| Diesel range organics, DRO (C10 - 24) | 10 | mg/kg | MCERTS | 10 | ug/l | None |
| Diesel range organics, DRO (C25 - 28) | 10 | mg/kg | None | 10 | ug/l | None |
| Lube oil range organics, LORO (C24 - 40) | 10 | mg/kg | MCERTS | 10 | ug/l | None |

Note: This information is for our CONSETT laboratory. For specific details for our LENHAM Laboratory please contact 01622850410 or enquiries@dets.co.uk.

Speciated TPH (Criteria Working Group - CWG)

Soils

Waters

| | LOD | Units | Accreditation | LOD | Units | Accreditation |
|-------------------------------|------|-------|---------------|-----|-------|---------------|
| TPH Aliphatic >C5-6 | 0.01 | mg/kg | None | 1 | µg/l | 17025 |
| TPH Aliphatic >C6-8 | 0.01 | mg/kg | None | 1 | µg/l | 17025 |
| TPH Aliphatic >C8-10 | 0.01 | mg/kg | None | 1 | µg/l | 17025 |
| TPH Aliphatic >C10-12 | 1.5 | mg/kg | MCERTS | 1 | µg/l | none |
| TPH Aliphatic >C12-16 | 1.2 | mg/kg | MCERTS | 1 | µg/l | none |
| TPH Aliphatic >C16-21 | 1.5 | mg/kg | MCERTS | 1 | µg/l | none |
| TPH Aliphatic >C21-35 | 3.4 | mg/kg | MCERTS | 1 | µg/l | none |
| TPH Aliphatic >C35-44 | 3.4 | mg/kg | None | - | - | - |
| TPH Aromatic >EC5-7 (Benzene) | 0.01 | mg/kg | None | 1 | µg/l | 17025 |
| TPH Aromatic >EC7-8 (Toluene) | 0.01 | mg/kg | None | 1 | µg/l | 17025 |
| TPH Aromatic >EC8-10 | 0.01 | mg/kg | None | 1 | µg/l | 17025 |
| TPH Aromatic >EC10-12 | 0.9 | mg/kg | MCERTS | 1 | µg/l | none |
| TPH Aromatic >EC12-16 | 0.5 | mg/kg | MCERTS | 1 | µg/l | none |
| TPH Aromatic >EC16-21 | 0.6 | mg/kg | MCERTS | 1 | µg/l | none |
| TPH Aromatic >EC21-35 | 1.4 | mg/kg | MCERTS | 1 | µg/l | none |
| TPH Aromatic >EC35-44 | 1.4 | mg/kg | None | - | - | - |
| Benzene | - | - | - | 1 | µg/l | 17025 |
| Toluene | - | - | - | 1 | µg/l | 17025 |

Note: This information is for our CONSETT laboratory. For specific details for our LENHAM Laboratory please contact 01622850410 or enquiries@dets.co.uk.



For the most meaningful results, it is crucial that samples are taken in the correct containers with no headspace, and are kept cold during transportation and storage prior to analysis.

Samples are analysed by headspace GCMS and run in full scan mode. Although the 64 target compounds are accurately quantified, additional compounds can be identified as TICs (tentatively identified compounds) - these must be requested when submitting samples for analysis.

Samples should be collected in 40 ml glass vials (waters) or 60g glass jars (soils), in duplicate, with no headspace.

Soils

| | Method | LOD | Units | Accreditation |
|-------------------------------|--------|------|-------|---------------|
| - o-Xylene | GCMS | 0.01 | mg/kg | 17025 |
| - Tetrachloroethylene | GCMS | 0.01 | mg/kg | 17025 |
| - 1,2,3-trichlorobenzene | GCMS | 0.01 | mg/kg | 17025 |
| - cis-1,3-dichloropropene | GCMS | 0.01 | mg/kg | 17025 |
| - m+p-Xylene | GCMS | 0.01 | mg/kg | 17025 |
| - 1,3-dichloropropane | GCMS | 0.01 | mg/kg | 17025 |
| - Chlorobenzene | GCMS | 0.01 | mg/kg | 17025 |
| - 1,3,5-trimethylbenzene | GCMS | 0.01 | mg/kg | 17025 |
| - 1,2-dibromoethane | GCMS | 0.01 | mg/kg | 17025 |
| - Dibromochloromethane | GCMS | 0.01 | mg/kg | 17025 |
| - Hexachlorobutadiene | GCMS | 0.01 | mg/kg | 17025 |
| - 1,1,2-trichloroethane | GCMS | 0.01 | mg/kg | 17025 |
| - trans-1,3-dichloropropene | GCMS | 0.01 | mg/kg | 17025 |
| - Toluene | GCMS | 0.01 | mg/kg | 17025 |
| - Bromodichloromethane | GCMS | 0.01 | mg/kg | 17025 |
| - Dibromomethane | GCMS | 0.01 | mg/kg | 17025 |
| - n-propylbenzene | GCMS | 0.01 | mg/kg | 17025 |
| - 2-chlorotoluene | GCMS | 0.01 | mg/kg | 17025 |
| - 1,2-dichloropropane | GCMS | 0.01 | mg/kg | 17025 |
| - Trichloroethylene | GCMS | 0.01 | mg/kg | 17025 |
| - 1,2-dichloroethane | GCMS | 0.01 | mg/kg | 17025 |
| - 1,1-dichloropropene | GCMS | 0.01 | mg/kg | 17025 |
| - Carbon tetrachloride | GCMS | 0.01 | mg/kg | 17025 |
| - Naphthalene | GCMS | 0.01 | mg/kg | 17025 |
| - Benzene | GCMS | 0.01 | mg/kg | 17025 |
| - n-butylbenzene | GCMS | 0.01 | mg/kg | 17025 |
| - 1,2-dichlorobenzene | GCMS | 0.01 | mg/kg | 17025 |
| - 1,2-dibromo-3-chloropropane | GCMS | 0.01 | mg/kg | 17025 |
| - 1,2,4-trichlorobenzene | GCMS | 0.01 | mg/kg | 17025 |
| - Tert-butylbenzene | GCMS | 0.01 | mg/kg | 17025 |
| - 1,2,4-trimethylbenzene | GCMS | 0.01 | mg/kg | 17025 |

Note: This information is for our CONSETT laboratory .

For specific details for our LENHAM Laboratory please contact 01622850410 or enquiries@dets.co.uk .

Continued on next page . .

Soils

| | Method | LOD | Units | Accreditation |
|------------------------------|--------|------|-------|---------------|
| - sec-butylbenzene | GCMS | 0.01 | mg/kg | 17025 |
| - 1,4-dichlorobenzene | GCMS | 0.01 | mg/kg | 17025 |
| - 1,2,3-trichloropropane | GCMS | 0.01 | mg/kg | 17025 |
| - Bromobenzene | GCMS | 0.01 | mg/kg | 17025 |
| - Isopropylbenzene | GCMS | 0.01 | mg/kg | 17025 |
| - 4-chlorotoluene | GCMS | 0.01 | mg/kg | 17025 |
| - MBTE | GCMS | 0.01 | mg/kg | None |
| - Vinyl Chloride | GCMS | 0.01 | mg/kg | 17025 |
| - Cis-1,2-dichloroethylene | GCMS | 0.01 | mg/kg | 17025 |
| - p-isopropyltoluene | GCMS | 0.01 | mg/kg | 17025 |
| - 1,3-dichlorobenzene | GCMS | 0.01 | mg/kg | 17025 |
| - 1,1,1,2-tetrachloroethane | GCMS | 0.01 | mg/kg | 17025 |
| - Ethylbenzene | GCMS | 0.01 | mg/kg | 17025 |
| - 1,1,1-trichloroethane | GCMS | 0.01 | mg/kg | 17025 |
| - Chloroform | GCMS | 0.01 | mg/kg | 17025 |
| - 1,1-dichloroethane | GCMS | 0.01 | mg/kg | 17025 |
| - Trans-1,2-dichloroethylene | GCMS | 0.01 | mg/kg | 17025 |
| - 2,2-dichloropropane | GCMS | 0.01 | mg/kg | 17025 |
| - Bromoform | GCMS | 0.01 | mg/kg | 17025 |
| - Bromochloromethane | GCMS | 0.01 | mg/kg | 17025 |
| - Styrene | GCMS | 0.01 | mg/kg | None |
| - 1,1 Dichloroethylene | GCMS | 0.01 | mg/kg | 17025 |

Note: This information is for our CONSETT laboratory .

For specific details for our LENHAM Laboratory please contact 01622850410 or enquiries@dets.co.uk.



Waters

| | Method | LOD | Units | Accreditation |
|-------------------------------|--------|-----|-------|---------------|
| - 1,2,3-trichlorobenzene | GCMS | 1 | ug/l | 17025 |
| - Carbon tetrachloride | GCMS | 1 | ug/l | 17025 |
| - 1,1-dichloropropene | GCMS | 1 | ug/l | 17025 |
| - Ethylbenzene | GCMS | 1 | ug/l | 17025 |
| - 1,1,1,2-tetrachloroethane | GCMS | 1 | ug/l | 17025 |
| - 1,3-dichlorobenzene | GCMS | 2 | ug/l | 17025 |
| - p-isopropyltoluene | GCMS | 1 | ug/l | 17025 |
| - 2,2-dichloropropane | GCMS | 2 | ug/l | 17025 |
| - Dichlorodifluoromethane | GCMS | 1 | ug/l | 17025 |
| - Chloromethane | GCMS | 1 | ug/l | 17025 |
| - Bromomethane | GCMS | 1 | ug/l | 17025 |
| - Chloroethane | GCMS | 1 | ug/l | 17025 |
| - Trichlorofluoromethane | GCMS | 1 | ug/l | None |
| - 1,1,2,2-tetrachloroethane | GCMS | 1 | ug/l | 17025 |
| - Vinyl Chloride | GCMS | 1 | ug/l | 17025 |
| - MTBE | GCMS | 1 | ug/l | None |
| - Naphthalene | GCMS | 1 | ug/l | 17025 |
| - Hexachlorobutadiene | GCMS | 1 | ug/l | 17025 |
| - 1,2,4-trichlorobenzene | GCMS | 1 | ug/l | 17025 |
| - 1,2-dibromo-3-chloropropane | GCMS | 1 | ug/l | 17025 |
| - 1,2-dichlorobenzene | GCMS | 1 | ug/l | 17025 |
| - n-butylbenzene | GCMS | 1 | ug/l | 17025 |
| - 1,4-dichlorobenzene | GCMS | 1 | ug/l | 17025 |
| - sec-butylbenzene | GCMS | 1 | ug/l | 17025 |
| - 1,2,4-trimethylbenzene | GCMS | 1 | ug/l | 17025 |
| - Tert-butylbenzene | GCMS | 1 | ug/l | 17025 |
| - 4-chlorotoluene | GCMS | 1 | ug/l | 17025 |
| - 1,3,5-trimethylbenzene | GCMS | 1 | ug/l | 17025 |
| - 2-chlorotoluene | GCMS | 1 | ug/l | 17025 |
| - n-propylbenzene | GCMS | 1 | ug/l | 17025 |
| - 1,2,3-trichloropropane | GCMS | 1 | ug/l | 17025 |
| - Bromobenzene | GCMS | 1 | ug/l | 17025 |
| - Isopropylbenzene | GCMS | 1 | ug/l | 17025 |
| - Bromoform | GCMS | 1 | ug/l | 17025 |
| - Styrene | GCMS | 1 | ug/l | 17025 |
| - o-Xylene | GCMS | 1 | ug/l | 17025 |
| - m+p-Xylene | GCMS | 2 | ug/l | 17025 |
| - Chlorobenzene | GCMS | 1 | ug/l | 17025 |
| - 1,2-dibromoethane | GCMS | 1 | ug/l | 17025 |

Note: This information is for our CONSETT laboratory .

For specific details for our LENHAM Laboratory please contact 01622850410 or enquiries@dets.co.uk .

Continued on next page . .

Waters

| | Method | LOD | Units | Accreditation |
|------------------------------|--------|-----|-------|---------------|
| - Dibromochloromethane | GCMS | 1 | ug/l | 17025 |
| - 1,3-dichloropropane | GCMS | 1 | ug/l | 17025 |
| - Tetrachloroethylene | GCMS | 1 | ug/l | 17025 |
| - 1,1,2-trichloroethane | GCMS | 1 | ug/l | 17025 |
| - trans-1,3-dichloropropene | GCMS | 1 | ug/l | 17025 |
| - Toluene | GCMS | 1 | ug/l | 17025 |
| - Trans-1,2-dichloroethylene | GCMS | 1 | ug/l | 17025 |
| - 1,1-dichloroethane | GCMS | 1 | ug/l | 17025 |
| - Cis-1,2-dichloroethylene | GCMS | 1 | ug/l | 17025 |
| - Bromochloromethane | GCMS | 4 | ug/l | 17025 |
| - Chloroform | GCMS | 1 | ug/l | 17025 |
| - 1,1-dichloroethylene | GCMS | 1 | ug/l | 17025 |
| - 1,1,1-trichloroethane | GCMS | 1 | ug/l | 17025 |
| - Benzene | GCMS | 1 | ug/l | 17025 |
| - 1,2-dichloroethane | GCMS | 1 | ug/l | 17025 |
| - Trichloroethylene | GCMS | 1 | ug/l | None |
| - Dibromomethane | GCMS | 1 | ug/l | 17025 |
| - 1,2-dichloropropane | GCMS | 1 | ug/l | 17025 |
| - Bromodichloromethane | GCMS | 4 | ug/l | 17025 |
| - cis-1,3-dichloropropene | GCMS | 1 | ug/l | 17025 |

Note: This information is for our CONSETT laboratory .

For specific details for our LENHAM Laboratory please contact 01622850410 or enquiries@dets.co.uk.



SVOCs are determined using GCMS, but a solvent extraction of the sample is required prior to analysis. As for VOCs, it is possible to request TICs in addition to the target list.

Soils

Waters

| | Method | LOD | Units | Accreditation | Method | LOD | Units | Accreditation |
|-------------------------------|--------|-----|-------|---------------|--------|-----|-------|---------------|
| - 2,4,5-Trichlorophenol | GCMS | 0.1 | mg/kg | None | GCMS | 1 | ug/l | None |
| - 2,3,4,6-Tetrachlorophenol | GCMS | 0.1 | mg/kg | None | GCMS | 1 | ug/l | None |
| - 4-Bromophenylphenylether | GCMS | 0.1 | mg/kg | 17025 | GCMS | 1 | ug/l | None |
| - Azobenzene | GCMS | 0.1 | mg/kg | 17025 | GCMS | 1 | ug/l | None |
| - Dibenzofuran | GCMS | 0.1 | mg/kg | 17025 | GCMS | 1 | ug/l | None |
| - Acenaphthene | GCMS | 0.1 | mg/kg | 17025 | GCMS | 1 | ug/l | None |
| - 3-Nitroaniline | GCMS | 0.1 | mg/kg | None | GCMS | 1 | ug/l | None |
| - 1,2-Dinitrobenzene | GCMS | 0.1 | mg/kg | None | GCMS | 1 | ug/l | None |
| - 2,4-Dinitrotoluene | GCMS | 0.1 | mg/kg | None | GCMS | 1 | ug/l | None |
| - 4-Nitrophenol | GCMS | 0.1 | mg/kg | None | GCMS | 1 | ug/l | None |
| - 2,3,5,6-Tetrachlorophenol | GCMS | 0.1 | mg/kg | None | GCMS | 1 | ug/l | None |
| - 2,4,6-Trichlorophenol | GCMS | 0.1 | mg/kg | 17025 | GCMS | 1 | ug/l | None |
| - Hexachlorocyclopentadiene | GCMS | 0.1 | mg/kg | None | GCMS | 1 | ug/l | None |
| - 2-Methylnaphthalene | GCMS | 0.1 | mg/kg | 17025 | GCMS | 1 | ug/l | None |
| - 4-Chloro-3-methylphenol | GCMS | 0.1 | mg/kg | 17025 | GCMS | 1 | ug/l | None |
| - Naphthalene | GCMS | 0.1 | mg/kg | 17025 | GCMS | 1 | ug/l | None |
| - Hexachlorobenzene | GCMS | 0.1 | mg/kg | 17025 | GCMS | 1 | ug/l | None |
| - 1,2,4-Trichlorobenzene | GCMS | 0.1 | mg/kg | 17025 | GCMS | 1 | ug/l | None |
| - Aniline | GCMS | 0.1 | mg/kg | None | GCMS | 1 | ug/l | None |
| - Bis(2-chloroisopropyl)ether | GCMS | 0.1 | mg/kg | 17025 | GCMS | 1 | ug/l | None |
| - Phenol | GCMS | 0.1 | mg/kg | 17025 | GCMS | 1 | ug/l | None |
| - 2-Chlorophenol | GCMS | 0.1 | mg/kg | 17025 | GCMS | 1 | ug/l | None |
| - Benzyl Alcohol | GCMS | 0.1 | mg/kg | 17025 | GCMS | 1 | ug/l | None |
| - 2-Methylphenol | GCMS | 0.1 | mg/kg | 17025 | GCMS | 1 | ug/l | None |
| - 3&4-Methylphenol | GCMS | 0.1 | mg/kg | 17025 | GCMS | 1 | ug/l | None |
| - 2,4-Dimethylphenol | GCMS | 0.1 | mg/kg | 17025 | GCMS | 1 | ug/l | None |
| - Bis-(dichloroethoxy)methane | GCMS | 0.1 | mg/kg | 17025 | GCMS | 1 | ug/l | None |
| - 2,4-Dichlorophenol | GCMS | 0.1 | mg/kg | 17025 | GCMS | 1 | ug/l | None |
| - Carbazole | GCMS | 0.1 | mg/kg | None | GCMS | 1 | ug/l | None |
| - Benzo(ghi)perylene | GCMS | 0.1 | mg/kg | 17025 | GCMS | 1 | ug/l | None |
| - Dibenzo(ah)anthracene | GCMS | 0.1 | mg/kg | 17025 | GCMS | 1 | ug/l | None |
| - Indeno(123cd)pyrene | GCMS | 0.1 | mg/kg | 17025 | GCMS | 1 | ug/l | None |

Note: This information is for our CONSETT laboratory. For specific details for our LENHAM Laboratory please contact 01622850410 or enquiries@dets.co.uk.

Continued on next page . .

Soils

Waters

| | Method | LOD | Units | Accreditation | Method | LOD | Units | Accreditation |
|------------------------------|--------|-----|-------|---------------|--------|-----|-------|---------------|
| - Benzo(a)pyrene | GCMS | 0.1 | mg/kg | 17025 | GCMS | 1 | ug/l | n/a |
| - Benzo(k)fluoranthene | GCMS | 0.1 | mg/kg | 17025 | GCMS | 1 | ug/l | n/a |
| - Benzo(b)fluoranthene | GCMS | 0.1 | mg/kg | 17025 | GCMS | 1 | ug/l | n/a |
| - Di-n-octylphthalate | GCMS | 0.1 | mg/kg | 17025 | GCMS | 1 | ug/l | n/a |
| - Chrysene | GCMS | 0.1 | mg/kg | 17025 | GCMS | 1 | ug/l | n/a |
| - Benzo(a)anthracene | GCMS | 0.1 | mg/kg | 17025 | GCMS | 1 | ug/l | n/a |
| - Pyrene | GCMS | 0.1 | mg/kg | 17025 | GCMS | 1 | ug/l | n/a |
| - Fluoranthene | GCMS | 0.1 | mg/kg | 17025 | GCMS | 1 | ug/l | n/a |
| - Bis(2-ethylhexyl)phthalate | GCMS | 0.1 | mg/kg | 17025 | GCMS | 1 | ug/l | n/a |
| - Butylbenzylphthalate | GCMS | 0.1 | mg/kg | None | GCMS | 1 | ug/l | n/a |
| - Di-n-butylphthalate | GCMS | 0.1 | mg/kg | 17025 | GCMS | 1 | ug/l | n/a |
| - Anthracene | GCMS | 0.1 | mg/kg | 17025 | GCMS | 1 | ug/l | n/a |
| - 2-Methyl-4,6-Dinitrophenol | GCMS | 0.1 | mg/kg | None | GCMS | 1 | ug/l | n/a |
| - 4-Nitroaniline | GCMS | 0.1 | mg/kg | None | GCMS | 1 | ug/l | n/a |
| - 4-Chlorophenylphenylether | GCMS | 0.1 | mg/kg | None | GCMS | 1 | ug/l | n/a |
| - Fluorene | GCMS | 0.1 | mg/kg | 17025 | GCMS | 1 | ug/l | n/a |
| - Diethylphthalate | GCMS | 0.1 | mg/kg | 17025 | GCMS | 1 | ug/l | n/a |
| - Diphenylamine | GCMS | 0.1 | mg/kg | 17025 | GCMS | 1 | ug/l | n/a |
| - 2-Chloronaphthalene | GCMS | 0.1 | mg/kg | 17025 | GCMS | 1 | ug/l | n/a |
| - 2-Nitroaniline | GCMS | 0.1 | mg/kg | None | GCMS | 1 | ug/l | n/a |
| - Pentachlorophenol | GCMS | 0.1 | mg/kg | None | GCMS | 1 | ug/l | n/a |
| - 2,6-Dinitrotoluene | GCMS | 0.1 | mg/kg | 17025 | GCMS | 1 | ug/l | n/a |
| - Acenaphthylene | GCMS | 0.1 | mg/kg | 17025 | GCMS | 1 | ug/l | n/a |
| - 1,3-Dinitrobenzene | GCMS | 0.1 | mg/kg | None | GCMS | 1 | ug/l | n/a |
| - Phenanthrene | GCMS | 0.1 | mg/kg | 17025 | GCMS | 1 | ug/l | n/a |
| - 1,4-Dinitrobenzene | GCMS | 0.1 | mg/kg | None | GCMS | 1 | ug/l | n/a |
| - Dimethylphthalate | GCMS | 0.1 | mg/kg | 17025 | GCMS | 1 | ug/l | n/a |

Note: This information is for our CONSETT laboratory. For specific details for our LENHAM Laboratory please contact 01622850410 or enquiries@dets.co.uk.

Soils

PAHs are some of the most commonly requested analyses, and are determined by solvent extraction, followed by GCMS.

Benzo(a)pyrene is a common marker compound requested by the regulators.

| | Method | LOD | Units | Accreditation |
|-------------------------|--------|------|-------|---------------|
| Acenaphthene | GCMS | 0.03 | mg/kg | MCERTS |
| Acenaphthylene | GCMS | 0.03 | mg/kg | MCERTS |
| Anthracene | GCMS | 0.03 | mg/kg | 17025 |
| Benzo(a)Anthracene | GCMS | 0.03 | mg/kg | MCERTS |
| Benzo(b)fluoranthene | GCMS | 0.03 | mg/kg | MCERTS |
| Benzo(k)fluoranthene | GCMS | 0.03 | mg/kg | MCERTS |
| Benzo(g,h,i)perylene | GCMS | 0.03 | mg/kg | MCERTS |
| Benzo(a)Pyrene | GCMS | 0.03 | mg/kg | MCERTS |
| Chrysene | GCMS | 0.03 | mg/kg | 17025 |
| Di-benzo(a,h)anthracene | GCMS | 0.03 | mg/kg | MCERTS |
| Fluoranthene | GCMS | 0.03 | mg/kg | MCERTS |
| Fluorene | GCMS | 0.03 | mg/kg | 17025 |
| Indeno(1,2,3-cd)pyrene | GCMS | 0.03 | mg/kg | MCERTS |
| Naphthalene | GCMS | 0.03 | mg/kg | MCERTS |
| Phenanthrene | GCMS | 0.03 | mg/kg | MCERTS |
| Pyrene | GCMS | 0.03 | mg/kg | MCERTS |

Note: This information is for our CONSETT laboratory. For specific details for our LENHAM Laboratory please contact 01622850410 or enquiries@dets.co.uk.

| | Method | LOD | Units | Accreditation |
|-------------------------|--------|-----|-------|---------------|
| Acenaphthene | GCFID | 0.1 | mg/kg | 17025 |
| Acenaphthylene | GCFID | 0.1 | mg/kg | 17025 |
| Anthracene | GCFID | 0.1 | mg/kg | 17025 |
| Benzo(a)Anthracene | GCFID | 0.1 | mg/kg | 17025 |
| Benzo(b)fluoranthene | GCFID | 0.1 | mg/kg | 17025 |
| Benzo(k)fluoranthene | GCFID | 0.1 | mg/kg | 17025 |
| Benzo(g,h,i)perylene | GCFID | 0.1 | mg/kg | 17025 |
| Benzo(a)Pyrene | GCFID | 0.1 | mg/kg | 17025 |
| Chrysene | GCFID | 0.1 | mg/kg | 17025 |
| Di-benzo(a,h)anthracene | GCFID | 0.1 | mg/kg | 17025 |
| Fluoranthene | GCFID | 0.1 | mg/kg | 17025 |
| Fluorene | GCFID | 0.1 | mg/kg | 17025 |
| Indeno(1,2,3-cd)pyrene | GCFID | 0.1 | mg/kg | 17025 |
| Naphthalene | GCFID | 0.1 | mg/kg | 17025 |
| Phenanthrene | GCFID | 0.1 | mg/kg | 17025 |
| Pyrene | GCFID | 0.1 | mg/kg | 17025 |

Note: This information is for our CONSETT laboratory. For specific details for our LENHAM Laboratory please contact 01622850410 or enquiries@dets.co.uk.

Waters

| | Method | LOD | Units | Accreditation |
|-----------------------|--------|------|-------|---------------|
| Acenaphthene | GCMS | 0.01 | µg/l | 17025 |
| Anthracene | GCMS | 0.01 | µg/l | 17025 |
| Acenaphthylene | GCMS | 0.01 | µg/l | 17025 |
| Benzo(a)anthracene | GCMS | 0.01 | µg/l | 17025 |
| Benzo(a)pyrene | GCMS | 0.01 | µg/l | 17025 |
| Benzo(b)fluoranthene | GCMS | 0.01 | µg/l | 17025 |
| Benzo(ghi)perylene | GCMS | 0.01 | µg/l | 17025 |
| Benzo(k)fluoranthene | GCMS | 0.01 | µg/l | 17025 |
| Chrysene | GCMS | 0.01 | µg/l | 17025 |
| Dibenzo(ah)anthracene | GCMS | 0.01 | µg/l | 17025 |
| Fluoranthene | GCMS | 0.01 | µg/l | 17025 |
| Fluorene | GCMS | 0.01 | µg/l | 17025 |
| Indeno(123cd)pyrene | GCMS | 0.01 | µg/l | 17025 |
| Naphthalene | GCMS | 0.05 | µg/l | 17025 |
| Phenanthrene | GCMS | 0.01 | µg/l | 17025 |
| Pyrene | GCMS | 0.01 | µg/l | 17025 |
| PAH (Total) | GCMS | 0.20 | µg/l | 17025 |

Note: This information is for our CONSETT laboratory. For specific details for our LENHAM Laboratory please contact 01622850410 or enquiries@dets.co.uk.



PCBs can be expressed as a total, but the individual congeners/suites are most commonly requested.

PCB 7 Congeners

Soils

Waters

| Determinand | Method | LOD | Units | Accreditation | Method | LOD | Units | Accreditation |
|-------------|--------|------|-------|---------------|--------|-----|-------|---------------|
| 101 | GCMS | 0.01 | mg/kg | MCERTS | GCMS | 0.3 | ug/l | 17025 |
| 118 | GCMS | 0.01 | mg/kg | MCERTS | GCMS | 0.6 | ug/l | 17025 |
| 138 | GCMS | 0.01 | mg/kg | MCERTS | GCMS | 0.2 | ug/l | 17025 |
| 153 | GCMS | 0.01 | mg/kg | MCERTS | GCMS | 0.2 | ug/l | 17025 |
| 180 | GCMS | 0.01 | mg/kg | MCERTS | GCMS | 0.2 | ug/l | 17025 |
| 28/31 | GCMS | 0.01 | mg/kg | MCERTS | GCMS | 0.3 | ug/l | 17025 |
| 52 | GCMS | 0.01 | mg/kg | MCERTS | GCMS | 0.2 | ug/l | 17025 |

Note: This information is for our CONSETT laboratory. For specific details for our LENHAM Laboratory please contact 01622850410 or enquiries@dets.co.uk.

PCB WHO 12 congeners (dioxin like PCBs)

Soils

Waters

| Determinand | Method | LOD | Units | Accreditation | Method | LOD | Units | Accreditation |
|-------------|--------|------|-------|---------------|--------|-----|-------|---------------|
| 77 | GCMS | 0.01 | mg/kg | None | GCMS | 0.3 | ug/l | 17025 |
| 81 | GCMS | 0.01 | mg/kg | None | GCMS | 0.2 | ug/l | 17025 |
| 105 | GCMS | 0.01 | mg/kg | None | GCMS | 0.2 | ug/l | 17025 |
| 114 | GCMS | 0.01 | mg/kg | None | GCMS | 0.3 | ug/l | 17025 |
| 118/123 | GCMS | 0.01 | mg/kg | None | GCMS | 0.6 | ug/l | 17025 |
| 126 | GCMS | 0.01 | mg/kg | None | GCMS | 0.5 | ug/l | 17025 |
| 156 | GCMS | 0.01 | mg/kg | None | GCMS | 0.3 | ug/l | 17025 |
| 157 | GCMS | 0.01 | mg/kg | None | GCMS | 0.2 | ug/l | 17025 |
| 167 | GCMS | 0.01 | mg/kg | None | GCMS | 0.3 | ug/l | 17025 |
| 169 | GCMS | 0.01 | mg/kg | None | GCMS | 0.2 | ug/l | 17025 |
| 189 | GCMS | 0.01 | mg/kg | None | GCMS | 0.3 | ug/l | 17025 |

Note: This information is for our CONSETT laboratory. For specific details for our LENHAM Laboratory please contact 01622850410 or enquiries@dets.co.uk.

PCB 7 congeners and PCB WHO 12 congeners (dioxin like)

Soils

Waters

| Determinand | Method | LOD | Units | Accreditation | Method | LOD | Units | Accreditation |
|-------------|--------|------|-------|---------------|--------|-----|-------|---------------|
| 101 | GCMS | 0.01 | mg/kg | MCERTS | GCMS | 0.3 | ug/l | 17025 |
| 118 | GCMS | 0.01 | mg/kg | MCERTS | GCMS | 0.6 | ug/l | 17025 |
| 138 | GCMS | 0.01 | mg/kg | MCERTS | GCMS | 0.2 | ug/l | 17025 |
| 153 | GCMS | 0.01 | mg/kg | MCERTS | GCMS | 0.2 | ug/l | 17025 |
| 180 | GCMS | 0.01 | mg/kg | MCERTS | GCMS | 0.2 | ug/l | 17025 |
| 28/31 | GCMS | 0.01 | mg/kg | MCERTS | GCMS | 0.3 | ug/l | 17025 |
| 52 | GCMS | 0.01 | mg/kg | MCERTS | GCMS | 0.2 | ug/l | 17025 |
| 77 | GCMS | 0.01 | mg/kg | None | GCMS | 0.3 | ug/l | 17025 |
| 81 | GCMS | 0.01 | mg/kg | None | GCMS | 0.2 | ug/l | 17025 |
| 105 | GCMS | 0.01 | mg/kg | None | GCMS | 0.2 | ug/l | 17025 |
| 114 | GCMS | 0.01 | mg/kg | None | GCMS | 0.3 | ug/l | 17025 |
| 118/123 | GCMS | 0.01 | mg/kg | None | GCMS | 0.6 | ug/l | 17025 |
| 126 | GCMS | 0.01 | mg/kg | None | GCMS | 0.5 | ug/l | 17025 |
| 156 | GCMS | 0.01 | mg/kg | None | GCMS | 0.3 | ug/l | 17025 |
| 157 | GCMS | 0.01 | mg/kg | None | GCMS | 0.2 | ug/l | 17025 |
| 167 | GCMS | 0.01 | mg/kg | None | GCMS | 0.3 | ug/l | 17025 |
| 169 | GCMS | 0.01 | mg/kg | None | GCMS | 0.2 | ug/l | 17025 |
| 189 | GCMS | 0.01 | mg/kg | None | GCMS | 0.3 | ug/l | 17025 |

Note: This information is for our CONSETT laboratory . For specific details for our LENHAM Laboratory please contact 01622850410 or enquiries@dets.co.uk.



Phenols are some of the most common contaminants in soils and waters, due to their high solubility, and can be extremely toxic both to plants and animals.

Phenols by GCMS

Soils

Waters

| | Method | LOD | Units | Accreditation | Method | LOD | Units | Accreditation |
|---------------------------|--------|------|-------|---------------|--------|-----|-------|---------------|
| Phenol | GCMS | 0.01 | mg/kg | None | GCMS | 0.1 | ug/l | None |
| 4-chloro-3-methylphenol | GCMS | 0.01 | mg/kg | None | GCMS | 0.1 | ug/l | None |
| 2-chlorophenol | GCMS | 0.01 | mg/kg | None | GCMS | 0.1 | ug/l | None |
| 2,4-dichlorophenol | GCMS | 0.01 | mg/kg | None | GCMS | 0.1 | ug/l | None |
| 2,4-dimethylphenol | GCMS | 0.01 | mg/kg | None | GCMS | 0.1 | ug/l | None |
| 2-nitrophenol | GCMS | 0.01 | mg/kg | None | GCMS | 0.1 | ug/l | None |
| 2,4-dinitrophenol | GCMS | 0.01 | mg/kg | None | GCMS | 0.1 | ug/l | None |
| Pentachlorophenol | GCMS | 0.01 | mg/kg | None | GCMS | 0.1 | ug/l | None |
| p-cresol | GCMS | 0.01 | mg/kg | None | GCMS | 0.1 | ug/l | None |
| 2,6-dimethylphenol | GCMS | 0.01 | mg/kg | None | GCMS | 0.1 | ug/l | None |
| 2,6-dichlorophenol | GCMS | 0.01 | mg/kg | None | GCMS | 0.1 | ug/l | None |
| 2,4,6-trichlorophenol | GCMS | 0.01 | mg/kg | None | GCMS | 0.1 | ug/l | None |
| 2,3,4,6-tetrachlorophenol | GCMS | 0.01 | mg/kg | None | GCMS | 0.1 | ug/l | None |
| 2-methylphenol | - | - | - | - | GCMS | 0.1 | ug/l | None |
| 3,5-dimethylphenol | - | - | - | - | GCMS | 0.1 | ug/l | None |
| 3+4-methylphenol | - | - | - | - | GCMS | 0.1 | ug/l | None |
| 4-chlorophenol | - | - | - | - | GCMS | 0.1 | ug/l | None |

Note: This information is for our CONSETT laboratory. For specific details for our LENHAM Laboratory please contact 01622850410 or enquiries@dets.co.uk.

Pesticides are very varied, with very different chemistries depending on their structure, so there is no overall screening test. Our combined pesticide suite selects some of the most common pesticides as an indicator. The method is solvent extraction followed by GCMS for specific target compounds.

Combined Pesticide Suite

Soils

Waters

| Determinand | Method | LOD | Units | Accreditation | LOD | Units | Accreditation |
|----------------------|--------|-----|-------|---------------|-----|-------|---------------|
| Aldrin | GCMS | 0.1 | mg/kg | None | 1 | ug/l | None |
| Alpha-BHC | GCMS | 0.1 | mg/kg | None | 1 | ug/l | None |
| Azinphos Methyl | GCMS | 0.1 | mg/kg | None | 1 | ug/l | None |
| Beta-BHC | GCMS | 0.1 | mg/kg | None | 1 | ug/l | None |
| Delta-BHC | GCMS | 0.1 | mg/kg | None | 1 | ug/l | None |
| Diazinon | GCMS | 0.1 | mg/kg | None | 1 | ug/l | None |
| Dichlorvos | GCMS | 0.1 | mg/kg | None | 1 | ug/l | None |
| Dieldrin | GCMS | 0.1 | mg/kg | None | 1 | ug/l | None |
| Endosulfan I | GCMS | 0.1 | mg/kg | None | 1 | ug/l | None |
| Endosulfan II | GCMS | 0.1 | mg/kg | None | 1 | ug/l | None |
| Endosulfan Sulphate | GCMS | 0.1 | mg/kg | None | 1 | ug/l | None |
| Endrin | GCMS | 0.1 | mg/kg | None | 1 | ug/l | None |
| Ethyl Parathion | GCMS | 0.1 | mg/kg | None | 1 | ug/l | None |
| Fenitrothion | GCMS | 0.1 | mg/kg | None | 1 | ug/l | None |
| Gamma-BHC(lindane) | GCMS | 0.1 | mg/kg | None | 1 | ug/l | None |
| Heptachlor | GCMS | 0.1 | mg/kg | None | 1 | ug/l | None |
| Heptachlor Epoxide | GCMS | 0.1 | mg/kg | None | 1 | ug/l | None |
| Malathion | GCMS | 0.1 | mg/kg | None | 1 | ug/l | None |
| Methyl Parathion | GCMS | 0.1 | mg/kg | None | 1 | ug/l | None |
| Methoxychlor (total) | GCMS | 0.1 | mg/kg | None | 1 | ug/l | None |
| Mevinphos | GCMS | 0.1 | mg/kg | None | 1 | ug/l | None |
| p, p' DDT | GCMS | 0.1 | mg/kg | None | 1 | ug/l | None |
| p,p'-DDE | GCMS | 0.1 | mg/kg | None | 1 | ug/l | None |
| p,p'-TDE(DDD) | GCMS | 0.1 | mg/kg | None | 1 | ug/l | None |

Note: This information is for our CONSETT laboratory. For specific details for our LENHAM Laboratory please contact 01622850410 or enquiries@dets.co.uk.

Organochlorine Pesticide Suite

Soils

Waters

| Determinand | Method | LOD | Units | Accreditation | LOD | Units | Accreditation |
|----------------------|--------|-----|-------|---------------|-----|-------|---------------|
| Aldrin | GCMS | 0.1 | mg/kg | None | 1 | ug/l | None |
| Alpha – BHC | GCMS | 0.1 | mg/kg | None | 1 | ug/l | None |
| Beta – BHC | GCMS | 0.1 | mg/kg | None | 1 | ug/l | None |
| Cis-chlordane | GCMS | 0.1 | mg/kg | None | 1 | ug/l | None |
| Dieldrin | GCMS | 0.1 | mg/kg | None | 1 | ug/l | None |
| Endosulfan Sulphate | GCMS | 0.1 | mg/kg | None | 1 | ug/l | None |
| Endosulphan I | GCMS | 0.1 | mg/kg | None | 1 | ug/l | None |
| Endosulphan II | GCMS | 0.1 | mg/kg | None | 1 | ug/l | None |
| Endrin | GCMS | 0.1 | mg/kg | None | 1 | ug/l | None |
| Gamma –BHC (Lindane) | GCMS | 0.1 | mg/kg | None | 1 | ug/l | None |
| Heptachlor | GCMS | 0.1 | mg/kg | None | 1 | ug/l | None |
| Heptachlor epoxide | GCMS | 0.1 | mg/kg | None | 1 | ug/l | None |
| Hexachlorobenzene | GCMS | 0.1 | mg/kg | None | 1 | ug/l | None |
| Isodrin | GCMS | 0.1 | mg/kg | None | 1 | ug/l | None |
| o, p'-DDE | GCMS | 0.1 | mg/kg | None | 1 | ug/l | None |
| o, p'-DDT | GCMS | 0.1 | mg/kg | None | 1 | ug/l | None |
| o, p'-Methoxychlor | GCMS | 0.1 | mg/kg | None | 1 | ug/l | None |
| o,p'-TDE (DDD) | GCMS | 0.1 | mg/kg | None | 1 | ug/l | None |
| p, p' DDT | GCMS | 0.1 | mg/kg | None | 1 | ug/l | None |
| p, p'-Methoxychlor | GCMS | 0.1 | mg/kg | None | 1 | ug/l | None |
| p,p'-DDE | GCMS | 0.1 | mg/kg | None | 1 | ug/l | None |
| p,p'-TDE(DDD) | GCMS | 0.1 | mg/kg | None | 1 | ug/l | None |
| Permethrin | GCMS | 0.1 | mg/kg | None | 1 | ug/l | None |
| Quintozene (PCNB) | GCMS | 0.1 | mg/kg | None | 1 | ug/l | None |
| Tecnazene | GCMS | 0.1 | mg/kg | None | 1 | ug/l | None |
| Trans-Chlordane | GCMS | 0.1 | mg/kg | None | 1 | ug/l | None |
| Triallate | GCMS | 0.1 | mg/kg | None | 1 | ug/l | None |
| Trifluralin | GCMS | 0.1 | mg/kg | None | 1 | ug/l | None |

Note: This information is for our CONSETT laboratory . For specific details for our LENHAM Laboratory please contact 01622850410 or enquiries@dets.co.uk.

Organophosphorous Pesticide Suite

Soils

Waters

| Determinand | LOD | Units | Accreditation | LOD | Units | Accreditation |
|-----------------------------|-----|-------|---------------|-----|-------|---------------|
| Dichlorvos | 0.1 | mg/kg | None | 1 | ug/l | None |
| Azinphos ethyl | 0.1 | mg/kg | None | 1 | ug/l | None |
| Azinphos methyl | 0.1 | mg/kg | None | 1 | ug/l | None |
| Carbophenothion | 0.1 | mg/kg | None | 1 | ug/l | None |
| Chlorfenvinphos | 0.1 | mg/kg | None | 1 | ug/l | None |
| Chlorpyrifos | 0.1 | mg/kg | None | 1 | ug/l | None |
| Diazinon (Dimpylate) | 0.1 | mg/kg | None | 1 | ug/l | None |
| Dimethoate | 0.1 | mg/kg | None | 1 | ug/l | None |
| Ethyl Parathion (Parathion) | 0.1 | mg/kg | None | 1 | ug/l | None |
| Fenitrothion | 0.1 | mg/kg | None | 1 | ug/l | None |
| Fenthion | 0.1 | mg/kg | None | 1 | ug/l | None |
| Malathion | 0.1 | mg/kg | None | 1 | ug/l | None |
| Methyl Parathion | 0.1 | mg/kg | None | 1 | ug/l | None |
| Mevinphos | 0.1 | mg/kg | None | 1 | ug/l | None |
| Phosalone | 0.1 | mg/kg | None | 1 | ug/l | None |
| Propetamphos | 0.1 | mg/kg | None | 1 | ug/l | None |
| Triazophos | 0.1 | mg/kg | None | 1 | ug/l | None |

Note: This information is for our CONSETT laboratory . For specific details for our LENHAM Laboratory please contact 01622850410 or enquiries@dets.co.uk.



Soils

Waters

| Determinand | Method | LOD | Units | Accreditation | LOD | Units | Accreditation |
|-------------------|---------|-----|-------|---------------|------|-------|---------------|
| Benazolin | LCMS MS | 0.1 | mg/kg | None | 0.02 | ug/l | None |
| Bentazone | LCMS MS | 0.1 | mg/kg | None | 0.02 | ug/l | 17025 |
| Bromoxynil | LCMS MS | 0.1 | mg/kg | None | 0.02 | ug/l | 17025 |
| Clopyralid | LCMS MS | 0.1 | mg/kg | None | 0.02 | ug/l | 17025 |
| 2,4 – D | LCMS MS | 0.1 | mg/kg | None | 0.02 | ug/l | None |
| 2,4 –DB | LCMS MS | 0.1 | mg/kg | None | 0.02 | ug/l | 17025 |
| Dicamba | LCMS MS | 0.1 | mg/kg | None | 0.02 | ug/l | 17025 |
| Dichloroprop | LCMS MS | 0.1 | mg/kg | None | 0.02 | ug/l | 17025 |
| Fenoprop | LCMS MS | 0.1 | mg/kg | None | 0.02 | ug/l | 17025 |
| loxynil | LCMS MS | 0.1 | mg/kg | None | 0.02 | ug/l | 17025 |
| MCPA | LCMS MS | 0.1 | mg/kg | None | 0.02 | ug/l | 17025 |
| MCPB | LCMS MS | 0.1 | mg/kg | None | 0.02 | ug/l | 17025 |
| Mecoprop | LCMS MS | 0.1 | mg/kg | None | 0.02 | ug/l | 17025 |
| Picloram | LCMS MS | 0.1 | mg/kg | None | 0.02 | ug/l | None |
| Pentachlorophenol | LCMS MS | 0.1 | mg/kg | None | 0.02 | ug/l | None |
| 2,4,5 – T | LCMS MS | 0.1 | mg/kg | None | 0.02 | ug/l | 17025 |
| 2,3,6 – TBA | LCMS MS | 0.1 | mg/kg | None | 0.02 | ug/l | 17025 |
| Triclopyr | LCMS MS | 0.1 | mg/kg | None | 0.02 | ug/l | None |

Note: This information is for our CONSETT laboratory. For specific details for our LENHAM Laboratory please contact 01622850410 or enquiries@dets.co.uk.



Due to the large number of these sites in the UK, a specific suite of testing was devised, in order to cover the most likely contaminants.

Soils

| Determinand | LOD | Units | Accreditation |
|---|--------------|----------|------------------|
| Free Cyanide | 0.1 | mg/kg | MCERTS |
| Total Cyanide | 0.1 | mg/kg | MCERTS |
| Complex Cyanide | 0.2 | mg/kg | None |
| Elemental Sulphur | 0.75 | mg/kg | MCERTS |
| Soluble sulphate (2:1 extract) | 10 | mg/l | MCERTS |
| Chloride (soluble) | 1 | mg/kg | 17025 |
| Ammoniacal Nitrogen as NH ₄ | 0.5 | mg/kg | MCERTS |
| As(0.2), Cd(0.1), Cr(0.15), Cu(0.2), Hg(0.05), Ni(1), Pb(0.3), Se(0.5), Zn(1), Co(0.7), Mo(0.4) | See Brackets | mg/kg | MCERTS |
| Speciated phenols by HPLC | Various | mg/kg | None |
| PAH 16 by GC-MS | 0.03 | mg/kg | MCERTS (13 only) |
| TPH CWG fractions | 0.01-3.4 | mg/kg | MCERTS |
| BTEX/MTBE by GC-MS | Various | mg/kg | MCERTS |
| pH | 1.0 | pH units | MCERTS |
| Loss On Ignition (LOI) | 0.01 | % | MCERTS |
| Natural Moisture Content | 0.1 | % | 17025 |
| % Stones | 1 | % | None |
| Hexavalent Chromium | 1 | mg/kg | None |

Note: This information is for our CONSETT laboratory. For specific details for our LENHAM Laboratory please contact 01622850410 or enquiries@dets.co.uk.

NG Soil Leachate

| Determinand | LOD | Units | Accreditation |
|---|--------------|----------|-------------------|
| CEN 10:1 leachate prep - default | | | None |
| Complex Cyanide - Low Level | 0.1 | ug/l | 17025 |
| Free cyanide - Low Level | 0.1 | ug/l | 17025 |
| Total cyanide - Low Level | 0.1 | ug/l | 17025 |
| Total Sulphur | 10 | mg/l | None |
| Sulphate | 0.1 | mg/l | 17025 |
| Chloride | 0.1 | mg/l | 17025 |
| Total Ammonium | 0.015 | mg/l | 17025 |
| Low Level As(0.16), Cd(0.02), Cr(0.25), Cu(0.3), Pb(0.09), Hg(0.001), Ni(0.5), Se(0.25), Zn(0.5), Co(0.16), Fe(5.5), Mo(1.1), V(0.6), | See brackets | ug/l | 17025 |
| Hexavalent Chromium - Low Level | 7 | ug/l | 17025 |
| Speciated phenols by HPLC - Resorcinol, Catechol, Phenol, Total Cresols, Total Xylenols, 1-naphthol, 2,3,5-trimethyl phenol | 0.1 | ug/l | None |
| PAH 16 by GC-MS | 0.01 | ug/l | 17025 |
| TPH CWG Fractions | See p .51 | ug/l | See p .51 |
| BTEX/MTBE by GC-MS Benzene (1), Toluene (1), Ethyl Benzene (1), m/p Xylene (2), o Xylene (1), MTBE (1) | See brackets | ug/l | 17025 except MTBE |
| pH | 1 | pH units | 17025 |
| Total Organic Carbon (1) and/or DOC (2) | See brackets | mg/l | 17025 |

Note: This information is for our CONSETT laboratory. For specific details for our LENHAM Laboratory please contact 01622850410 or enquiries@dets.co.uk.

NG Water

| Determinand | LOD | Units | Accreditation |
|---|--------------|----------|--------------------|
| Complex Cyanide - Low Level | 0.1 | ug/l | 17025 |
| Free cyanide - Low Level | 0.1 | ug/l | 17025 |
| Total cyanide - Low Level | 0.1 | ug/l | 17025 |
| Total Sulphur | 10 | mg/l | None |
| Sulphate | 0.1 | mg/l | 17025 |
| Sulphide | 10 | ug/l | 17025 |
| Chloride | 0.1 | mg/l | 17025 |
| Total Ammonium | 0.15 | mg/l | 17025 |
| Low Level As(0.16), Cd(0.02), Cr(0.25), Cu(0.3), Pb(0.09), Hg(0.001), Ni(0.5), Se(0.25), Zn(0.5), Co(0.16), Fe(5.5), Mo(1.1), V(0.6), B(12) | See brackets | ug/l | 17025 except Boron |
| Speciated phenols by HPLC - Resorcinol, Catechol, Phenol, Total Cresols, Total Xylenols, 1-naphthol, 2,3,5-trimethyl phenol | 0.1 | ug/l | None |
| PAH 16 by GC-MS | 0.01 | ug/l | 17025 |
| TPH CWG fractions | See p .51 | ug/l | See p .51 |
| BTEX/MTBE by GC-MS (Benzene (1), Toluene (1), Ethyl Benzene (1), m/p Xylene (2), o Xylene (1), MTBE (1)) | See brackets | ug/l | 17025 except MTBE |
| pH | 1 | pH units | 17025 |
| Total Organic Carbon (1) and/or DOC (2) | See brackets | mg/l | 17025 |
| Total Suspended Solids | 5 | mg/l | 17025 |
| Electrical Conductivity | 1 | uS/cm | 17025 |
| Hexavalent Chromium - Low Level | 7 | ug/l | 17025 |

Note: This information is for our CONSETT laboratory . For specific details for our LENHAM Laboratory please contact 01622850410 or enquiries@dets.co.uk .

NG MNA

| Determinand | LOD | Units | Accreditation |
|---------------------------------------|------|-------|---------------|
| Total Alkalinity as CaCO ₃ | 10 | mg/l | 17025 |
| Dissolved methane | 25 | ug/l | None |
| Nitrate as NO ₃ | 0.1 | mg/l | 17025 |
| Nitrite as NO ₂ | 0.1 | mg/l | 17025 |
| Manganese II | 100 | ug/l | None |
| Iron II | 100 | ug/l | None |
| Iron III | 100 | ug/l | None |
| Dissolved CO ₂ | 0.01 | mg/l | None |

Note: This information is for our CONSETT laboratory . For specific details for our LENHAM Laboratory please contact 01622850410 or enquiries@dets.co.uk .

Gasholder Water Suite

| Determinand | LOD | Units | Accreditation |
|--|------------------------|----------|---------------|
| True Colour | 1 | Hazen | None |
| Dissolved Methane | 25 | ug/l | None |
| Dissolved Oxygen | 2 | mg/l | None |
| COD | 10 | mg/l | 17025 |
| BOD | 1 | mg/l | 17025 |
| Total Suspended Solids | 5 | mg/l | 17025 |
| pH | 1 | pH units | 17025 |
| Sulphide | 10 | mg/l | 17025 |
| Sulphate | 0.1 | mg/l | 17025 |
| Nitrate | 0.1 | mg/l | 17025 |
| Phenols | 0.1 | ug/l | None |
| Ammonia | 0.015 | mg/l | None |
| Ammoniacal Nitrogen | 0.015 | mg/l | 17025 |
| Fats, Oils and Grease | 1 | mg/l | 17025 |
| Metals : Fe(5.5), Zn(1.3), Cu(0.4), Hg(0.01), Cd(0.03), Cr(0.25), As(0.16), Pb(0.09) | see brackets | mg/l | 17025 |
| Total Cyanide | 40 | ug/l | 17025 |
| BTEX | 1 m/p Xylene (2) | mg/l | 17025 |
| TOC | 1 | mg/l | 17025 |
| PAH USEPA 16 | 0.01 | ug/l | 17025 |
| TPH CWG Fractions | see below | ug/l | see below |
| TPH Aliphatic >C6-6 | 1 | ug/l | 17025 |
| TPH Aliphatic >C6-8 | 1 | ug/l | 17025 |
| TPH Aliphatic >C8-10 | 1 | ug/l | 17025 |
| TPH Aliphatic >C10-12 | 1 | ug/l | None |
| TPH Aliphatic >C12-16 | 1 | ug/l | None |
| TPH Aliphatic >C16-21 | 1 | ug/l | None |
| TPH Aliphatic >C21-35 | 1 | ug/l | None |
| TPH Aromatic >EC5-7 (Benzene) | 1 | ug/l | 17025 |
| TPH Aromatic >EC7-8 (Toluene) | 1 | ug/l | 17025 |
| TPH Aromatic >EC8-10 | 1 | ug/l | 17025 |
| TPH Aromatic >EC10-12 | 1 | ug/l | None |
| TPH Aromatic >EC12-16 | 1 | ug/l | None |
| TPH Aromatic >EC16-21 | 1 | ug/l | None |
| TPH Aromatic >EC21-35 | 1 | ug/l | None |

Note: This information is for our CONSETT laboratory. For specific details for our LENHAM Laboratory please contact 01622850410 or enquiries@dets.co.uk.

Watersystems will slowly regulate themselves, given the right conditions, and this suite is designed to monitor the progress.

| Determinand | LOD | Units | Accreditation |
|---------------------------------------|------|----------|---------------|
| Total Alkalinity as CaCO ₃ | 10 | mg/l | 17025 |
| Calcium | 0.09 | mg/l | 17025 |
| pH | 1 | pH units | 17025 |
| DOC | 2 | mg/l | 17025 |

Note: This information is for our CONSETT laboratory. For specific details for our LENHAM Laboratory please contact 01622850410 or enquiries@dets.co.uk.

MNA Suite

| Determinand | LOD | Units | Accreditation |
|---|------|-------|---------------|
| Total Alkalinity as CaCO ₃ | 10 | mg/l | 17025 |
| Dissolved methane | 25 | ug/l | None |
| Dissolved CO ₂ | 0.01 | mg/l | None |
| Manganese II | 100 | ug/l | None |
| Manganese IV | 100 | ug/l | None |
| Total Manganese | 0.22 | ug/l | 17025 |
| Iron II | 100 | ug/l | None |
| Iron III | 100 | ug/l | None |
| Total Iron | 5.5 | ug/l | 17025 |
| Nitrate as NO ₃ | 0.1 | mg/l | 17025 |
| Nitrite as NO ₂ (should be tested within 24 hours) | 0.1 | mg/l | 17025 |
| Sulphate | 0.1 | mg/l | 17025 |
| Sulphide | 10 | ug/l | 17025 |
| Total Inorganic Carbon | 2 | mg/l | None |
| Total Organic Carbon | 2 | mg/l | 17025 |

Note: This information is for our CONSETT laboratory. For specific details for our LENHAM Laboratory please contact 01622850410 or enquiries@dets.co.uk.

These suites list designated priority pollutants. Although the methods are accredited on waters, the actual leachate preparation is not accredited.

Landfill Directive List I

| | SOIL-LEACHATE | | |
|--|---------------|-------|---------------|
| Determinand | LOD | Units | Accreditation |
| VOC target list including BTEX/MTBE + TICs by GC-MS | 1 | ug/l | None |
| SVOC target list including PAHs, phenol and chlorinated phenols plus TICs by GC-MS | Various | ug/l | None |
| Organophosphorous pesticides (21 compounds) | Various | ug/l | None |
| Tributyltin, triphenyltin, dibutyltin | 1 | ug/l | None |
| Total Mercury | 0.01 | ug/l | None |
| Total Cadmium | 0.03 | ug/l | None |
| EPH (C8-40) (total or dissolved) by GC-FID including mineral oil by calculation | 10 | ug/l | None |
| Total cyanide | 40 | ug/l | 17025 |

Note: This information is for our CONSETT laboratory. For specific details for our LENHAM Laboratory please contact 01622850410 or enquiries@dets.co.uk.

Landfill Directive List II

| | SOIL-LEACHATE | | |
|---|---------------|-------------------------|---------------|
| Determinand | LOD | Units | Accreditation |
| Total As, Cr, Cu, Pb, Ni, Se, Zn, B, Ba, Co, Mo, Sb, V, Be, Tl, P, Sn, Ti, Te, Ag | Various | ug/l | None |
| Total Uranium | | currently subcontracted | |
| Organochlorine pesticides (33 compounds) | Various | ug/l | None |
| Acid Herbicides | Various | ug/l | None |
| Silica | 0.1 | mg/l | None |
| Ortho-Phosphate as PO4 | 0.01 | mg/l | 17025 |
| Fluoride | 0.1 | mg/l | None |
| Ammoniacal Nitrogen as NH3 | 0.015 | mg/l | 17025 |
| Nitrate as NO3 | 0.1 | mg/l | 17025 |

Note: This information is for our CONSETT laboratory. For specific details for our LENHAM Laboratory please contact 01622850410 or enquiries@dets.co.uk.

As with the Landfill Directive, these compounds are designated priority pollutants .

EA LIST 1 DANGEROUS SUBSTANCES (76/464/EEC)

Soils

Waters

| Determinand | LOD | Units | Accreditation | LOD | Units | Accreditation |
|---------------------------------------|------|-------|---------------|------|-------|---------------|
| Cadmium | 0.1 | mg/kg | MCERTS | 0.03 | ug/l | 17025 |
| Mercury | 0.05 | mg/kg | MCERTS | 0.01 | ug/l | 17025 |
| Carbon tetrachloride | 0.01 | mg/kg | 17025 | 1 | ug/l | 17025 |
| Chloroform | 0.01 | mg/kg | 17025 | 1 | ug/l | 17025 |
| 1,2,4 -trichlorobenzene | 0.01 | mg/kg | 17025 | 1 | ug/l | 17025 |
| 1,2,3 - trichlorobenzene | 0.01 | mg/kg | 17025 | 1 | ug/l | 17025 |
| Trichloroethene (trichloroethylene) | 0.01 | mg/kg | 17025 | 1 | ug/l | None |
| 1,2 -Dichloroethane | 0.01 | mg/kg | 17025 | 1 | ug/l | 17025 |
| Tetrachloroethene (perchloroethylene) | 0.01 | mg/kg | 17025 | 1 | ug/l | 17025 |
| Hexachlorobutadiene | 0.01 | mg/kg | 17025 | 1 | ug/l | 17025 |
| Hexachlorobenzene | 0.1 | mg/kg | 17025 | 1 | ug/l | None |
| Pentachlorophenol | 0.1 | mg/kg | None | 1 | ug/l | None |
| Aldrin | 0.1 | mg/kg | None | 1 | ug/l | None |
| Dieldrin | 0.1 | mg/kg | None | 1 | ug/l | None |
| Endrin | 0.1 | mg/kg | None | 1 | ug/l | None |
| o,p DDT | 0.1 | mg/kg | None | 1 | ug/l | None |
| p,p -DDT | 0.1 | mg/kg | None | 1 | ug/l | None |

Note: This information is for our CONSETT laboratory . For specific details for our LENHAM Laboratory please contact 01622850410 or enquiries@dets.co.uk.

EA LIST 1 DANGEROUS SUBSTANCES (76/464/EEC)

Soils

Waters

| Determinand | LOD | Units | Accreditation | LOD | Units | Accreditation |
|---------------------|-----|----------|---------------|------|----------|---------------|
| pH | 1 | pH Units | MCERTS | 1 | pH Units | 17025 |
| Triphenyltin | 1 | mg/kg | None | 1 | ug/l | None |
| Tributyltin | 1 | mg/kg | None | 1 | ug/l | None |
| 2,4 -D | 0.1 | mg/kg | None | 0.02 | ug/l | None |
| 2,4 -DB | 0.1 | mg/kg | None | 0.02 | ug/l | 17025 |
| Mecoprop | 0.1 | mg/kg | None | 0.02 | ug/l | 17025 |
| Bentazone | 0.1 | mg/kg | None | 0.02 | ug/l | 17025 |
| Linuron | 0.1 | mg/kg | None | 0.05 | ug/l | None |
| Trifluralin | 0.1 | mg/kg | None | 1 | ug/l | None |
| Endosulfan Sulphate | 0.1 | mg/kg | None | 1 | ug/l | None |
| Endosulphan I | 0.1 | mg/kg | None | 1 | ug/l | None |
| Endosulphan II | 0.1 | mg/kg | None | 1 | ug/l | None |

Note: This information is for our CONSETT laboratory . For specific details for our LENHAM Laboratory please contact 01622850410 or enquiries@dets.co.uk.

Continued on next page . . .

EA LIST 1 DANGEROUS SUBSTANCES (76/464/EEC)

Soils

Waters


| Determinand | LOD | Units | Accreditation | LOD | Units | Accreditation |
|-------------------------|------|-------|---------------|------|-------|---------------|
| Triazophos | 0.1 | mg/kg | None | 1 | ug/l | None |
| Dichlorvos | 0.1 | mg/kg | None | 1 | ug/l | None |
| Mevinphos | 0.1 | mg/kg | None | 1 | ug/l | None |
| Dimethoate | 0.1 | mg/kg | None | 1 | ug/l | None |
| Omethoate | 0.1 | mg/kg | None | 1 | ug/l | None |
| Fenitrothion | 0.1 | mg/kg | None | 1 | ug/l | None |
| Malathion | 0.1 | mg/kg | None | 1 | ug/l | None |
| Azinphos-methyl | 0.1 | mg/kg | None | 1 | ug/l | None |
| Triazines | 0.1 | mg/kg | None | 1 | ug/l | None |
| Atrazine | 0.1 | mg/kg | None | 1 | ug/l | None |
| Simazine | 0.1 | mg/kg | None | 1 | ug/l | None |
| Cyfluthrin | 0.1 | mg/kg | None | 1 | ug/l | None |
| Permethrin | 0.1 | mg/kg | None | 1 | ug/l | None |
| 2,4 -dichlorophenol | 0.1 | mg/kg | 17025 | 1 | ug/l | None |
| Naphthalene | 0.1 | mg/kg | 17025 | 1 | ug/l | None |
| 4-Chloro-3-methylphenol | 0.1 | mg/kg | 17025 | 1 | ug/l | None |
| 2 - Chlorophenol | 0.1 | mg/kg | 17025 | 1 | ug/l | None |
| Biphenyl | 0.1 | mg/kg | 17025 | 1 | ug/l | None |
| Benzene | 0.01 | mg/kg | 17025 | 1 | ug/l | 17025 |
| Toluene | 0.01 | mg/kg | 17025 | 1 | ug/l | 17025 |
| m,p- xylene | 0.01 | mg/kg | 17025 | 2 | ug/l | 17025 |
| o- xylene | 0.01 | mg/kg | 17025 | 1 | ug/l | 17025 |
| Arsenic | 0.2 | mg/kg | MCERTS | 0.16 | ug/l | 17025 |
| Boron | 0.2 | mg/kg | MCERTS | 12 | ug/l | 17025 |
| Chromium | 0.15 | mg/kg | MCERTS | 0.25 | ug/l | 17025 |
| Copper | 0.2 | mg/kg | MCERTS | 0.4 | ug/l | 17025 |
| Iron | 25 | mg/kg | MCERTS | 5.5 | ug/l | 17025 |
| Lead | 0.3 | mg/kg | MCERTS | 0.09 | ug/l | 17025 |
| Nickel | 1 | mg/kg | MCERTS | 0.5 | ug/l | 17025 |
| Vanadium | 0.8 | mg/kg | MCERTS | 0.6 | ug/l | 17025 |
| Zinc (Total) | 1 | mg/kg | MCERTS | 1.3 | ug/l | 17025 |
| Demeton | 0.1 | mg/kg | None | 1 | ug/l | None |

Note: This information is for our CONSETT laboratory . For specific details for our LENHAM Laboratory please contact 01622850410 or enquiries@dets.co.uk.

The environmental forensic approach integrates and leverages the expertise of environmental chemists from the laboratory in partnership with the consultant. Environmental forensic chemistry investigations typically utilize a multi-level tiered interpretive approach to support the identification of the source of the contamination that has been characterized by environmental sample analysis. To achieve a successful environmental forensic investigation, it will encompass an understanding of a variety of disciplines and the wisdom to select the most appropriate forensic tool(s) which are best suited for a particular set of facts. The tiered approach is broken down into the following Tiers:

- SUEZ Tier 1 Identification of spill release products in contaminated samples shapes the environmental forensic investigation and is the foundation in the development of multiple lines of evidence approach. At this level typically, both standard and specialised methodologies are employed to identify the contaminant, including total petroleum hydrocarbons, PAHs, PCBs, metals and chlorinated solvents.
- SUEZ Tier 2 At this level the investigator can now target the specific signature compounds unique to the contamination identified in Tier 1. Utilizing gas chromatography coupled with mass spectrometry techniques targeted families of compounds such as bicyclic sesquiterpanes, alkylated PAHs and terpane biomarkers which all can be used to determine the degree of similarity to source materials, degree of weathering and/or biodegradation.
- SUEZ Tier 3 Building on the evidence obtained from Tier 1 and Tier 2 investigational analysis, the data visualization can be represented in several different ways including chemical component histograms, double ratio cross plots and compound ratio analysis spider plots. All can be used to achieve the common objective of many forensic investigations translating environmental information to allocated responsibility for the contamination.

The successful execution requires an understanding of fate and transport as related to contaminant alterations, as well as the forensic tools employed for the investigation



DETS can offer a range of specialist analyses on request. Please contact us for more details.

